

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123021\
 Data File : VN070412.D
 Acq On : 30 Dec 2021 20:40
 Operator : JC/MD
 Sample : M5234-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P-1-(FB)

Quant Time: Dec 31 05:24:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	831616	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	1428903	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	1278380	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	410866	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	651662	54.500	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.000%	
35) Dibromofluoromethane	8.024	113	463455	53.606	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 107.220%	
50) Toluene-d8	10.443	98	1743344	49.361	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.720%	
62) 4-Bromofluorobenzene	12.733	95	604846	47.416	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 94.840%	
Target Compounds						
16) Acetone	4.296	43	26094	3.362	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123021\
Data File : VN070412.D
Acq On : 30 Dec 2021 20:40
Operator : JC/MD
Sample : M5234-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P-1-(FB)

Quant Time: Dec 31 05:24:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

