

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123021\
 Data File : VN070415.D
 Acq On : 30 Dec 2021 22:48
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Quant Time: Dec 31 05:25:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Amit Patel 01/03/2022
 Supervised By :mohammad ahmed 01/03/2022

01/03/2022
 Supervised By :mohammad
 ahmed

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	867356	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1452023	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	1395806	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	550436	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.440	65	626563	50.242	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.480%	
35) Dibromofluoromethane	8.021	113	452385	51.492	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.980%	
50) Toluene-d8	10.443	98	1803849	50.261	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.520%	
62) 4-Bromofluorobenzene	12.731	95	654952	50.527	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.060%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.073	85	513146	52.169	ug/l	100
3) Chloromethane	2.303	50	708685	52.965	ug/l	99
4) Vinyl Chloride	2.446	62	631289	51.820	ug/l	98
5) Bromomethane	2.837	94	326562	51.936	ug/l	99
6) Chloroethane	3.009	64	367467	50.088	ug/l	98
7) Trichlorofluoromethane	3.371	101	721441	49.202	ug/l	98
8) Diethyl Ether	3.824	74	346261	53.569	ug/l	80
9) 1,1,2-Trichlorotrifluo...	4.205	101	469894	52.395	ug/l	96
10) Methyl Iodide	4.417	142	664620	54.760	ug/l	99
11) Tert butyl alcohol	5.388	59	661721	318.676	ug/l	98
12) 1,1-Dichloroethene	4.175	96	466842	52.848	ug/l	91
13) Acrolein	4.039	56	418080	248.145	ug/l	99
14) Allyl chloride	4.838	41	1137408	56.791	ug/l #	84
15) Acrylonitrile	5.557	53	2049220	316.252	ug/l	99
16) Acetone	4.291	43	1809192	223.479	ug/l	100
17) Carbon Disulfide	4.524	76	1341620	55.639	ug/l	100
18) Methyl Acetate	4.859	43	1432430	61.168	ug/l	91
19) Methyl tert-butyl Ether	5.618	73	1777189	54.174	ug/l	94
20) Methylene Chloride	5.093	84	593746	53.009	ug/l #	85
21) trans-1,2-Dichloroethene	5.594	96	504765	53.968	ug/l	87
22) Diisopropyl ether	6.501	45	2220030	56.910	ug/l #	91
23) Vinyl Acetate	6.436	43	8792816	290.516	ug/l #	94
24) 1,1-Dichloroethane	6.388	63	1105944	54.677	ug/l	98
25) 2-Butanone	7.340	43	2948329	289.241	ug/l	91
26) 2,2-Dichloropropane	7.327	77	717099	44.940	ug/l	98
27) cis-1,2-Dichloroethene	7.327	96	602339	53.972	ug/l	87
28) Bromochloromethane	7.659	49	510128	51.570	ug/l #	75
29) Tetrahydrofuran	7.689	42	1953648	321.234	ug/l	87
30) Chloroform	7.820	83	1050779	53.849	ug/l	100
31) Cyclohexane	8.094	56	1030572	52.368	ug/l	88
32) 1,1,1-Trichloroethane	8.016	97	884154	53.738	ug/l	95
36) 1,1-Dichloropropene	8.223	75	770570	53.020	ug/l	96
37) Ethyl Acetate	7.415	43	1133661	60.178	ug/l	96
38) Carbon Tetrachloride	8.206	117	758159	54.631	ug/l	99
39) Methylcyclohexane	9.462	83	929535	52.829	ug/l	89
40) Benzene	8.461	78	2421888	54.355	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123021\
 Data File : VN070415.D
 Acq On : 30 Dec 2021 22:48
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Quant Time: Dec 31 05:25:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Amit Patel 01/03/2022
 Supervised By :mohammad ahmed 01/03/2022

01/03/2022
 Supervised By :mohammad
 ahmed

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
41) Methacrylonitrile	7.641	41	584024	61.981	ug/l	#	87
42) 1,2-Dichloroethane	8.531	62	875862	52.298	ug/l		99
43) Isopropyl Acetate	8.563	43	1725529	58.096	ug/l	#	94
44) Trichloroethene	9.218	130	550605	52.008	ug/l		92
45) 1,2-Dichloropropane	9.491	63	666580	55.564	ug/l		99
46) Dibromomethane	9.580	93	408834	54.805	ug/l		95
47) Bromodichloromethane	9.762	83	862788	58.248	ug/l		99
48) Methyl methacrylate	9.561	41	831451	58.689	ug/l		84
49) 1,4-Dioxane	9.572	88	275312	1327.563	ug/l	#	85
51) 4-Methyl-2-Pentanone	10.333	43	5994449	321.194	ug/l		96
52) Toluene	10.508	92	1487043	55.138	ug/l		98
53) t-1,3-Dichloropropene	10.719	75	903185	56.325	ug/l		99
54) cis-1,3-Dichloropropene	10.191	75	981844	56.602	ug/l		94
55) 1,1,2-Trichloroethane	10.899	97	596738	55.922	ug/l		96
56) Ethyl methacrylate	10.762	69	1050178	59.562	ug/l	#	89
57) 1,3-Dichloropropane	11.044	76	1054679	55.881	ug/l		100
58) 2-Chloroethyl Vinyl ether	10.041	63	1116742	248.112	ug/l		90
59) 2-Hexanone	11.087	43	4372428	321.267	ug/l		94
60) Dibromochloromethane	11.240	129	643732	61.830	ug/l		100
61) 1,2-Dibromoethane	11.347	107	616574	56.906	ug/l		99
64) Tetrachloroethene	10.980	164	490613	48.539	ug/l		97
65) Chlorobenzene	11.771	112	1547914	52.266	ug/l		99
66) 1,1,1,2-Tetrachloroethane	11.841	131	580221	54.879	ug/l		99
67) Ethyl Benzene	11.846	91	2914330	53.707	ug/l		97
68) m/p-Xylenes	11.953	106	2132038	107.940	ug/l		94
69) o-Xylene	12.280	106	1061964	54.053	ug/l		93
70) Styrene	12.294	104	1766081	56.796	ug/l		96
71) Bromoform	12.457	173	517568	65.960	ug/l	#	99
73) Isopropylbenzene	12.578	105	2785891	54.501	ug/l		98
74) N-amyl acetate	12.388	43	1292560	60.140	ug/l		95
75) 1,1,2,2-Tetrachloroethane	12.827	83	972333	59.114	ug/l		99
76) 1,2,3-Trichloropropane	12.881	75	949235m	57.800	ug/l		
77) Bromobenzene	12.860	156	636927	54.274	ug/l		76
78) n-propylbenzene	12.921	91	3151910	55.652	ug/l		96
79) 2-Chlorotoluene	13.007	91	1935069	53.250	ug/l		92
80) 1,3,5-Trimethylbenzene	13.061	105	2232675	54.506	ug/l		98
81) trans-1,4-Dichloro-2-b...	12.626	75	292142	64.845	ug/l		95
82) 4-Chlorotoluene	13.104	91	1831594	53.808	ug/l		93
83) tert-Butylbenzene	13.324	119	1972283	54.682	ug/l		95
84) 1,2,4-Trimethylbenzene	13.369	105	2199486	55.763	ug/l		97
85) sec-Butylbenzene	13.501	105	2715786	55.459	ug/l		97
86) p-Isopropyltoluene	13.616	119	2123989	55.216	ug/l		97
87) 1,3-Dichlorobenzene	13.616	146	1048880	53.391	ug/l		97
88) 1,4-Dichlorobenzene	13.694	146	1005907	51.908	ug/l		98
89) n-Butylbenzene	13.943	91	1721567	53.656	ug/l		96
90) Hexachloroethane	14.211	117	448850	67.167	ug/l		91
91) 1,2-Dichlorobenzene	13.989	146	1010883	52.489	ug/l		97
92) 1,2-Dibromo-3-Chloropr...	14.603	75	187477	63.313	ug/l		80
93) 1,2,4-Trichlorobenzene	15.263	180	501505	55.265	ug/l		98
94) Hexachlorobutadiene	15.370	225	278635	52.832	ug/l		99
95) Naphthalene	15.507	128	1539394	60.972	ug/l		100
96) 1,2,3-Trichlorobenzene	15.700	180	510073	58.619	ug/l		99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123021\
Data File : VN070415.D
Acq On : 30 Dec 2021 22:48
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 31 Sample Multiplier: 1

Quant Time: Dec 31 05:25:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022

01/03/2022
Supervised By :mohammad
ahmed

01/03/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed						



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.086 min Scan# 2273
Delta R.T. -0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:168 Resp: 867350
Ion Ratio Lower Upper
168 100
99 59.3 41.3 61.9



Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022






01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#2
Dichlorodifluoromethane
Concen: 52.169 ug/l
RT: 2.073 min Scan# 31
Delta R.T. -0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 85 Resp: 513146
Ion Ratio Lower Upper
85 100
87 32.7 16.3 48.9





#3
Chloromethane
Concen: 52.965 ug/l
RT: 2.303 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 50 Resp: 70868
Ion Ratio Lower Upper
50 100
52 32.1 25.4 38.0


Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022






01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#4
Vinyl Chloride
Concen: 51.820 ug/l
RT: 2.446 min Scan# 170
Delta R.T. -0.002 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 62 Resp: 631289
Ion Ratio Lower Upper
62 100
64 31.1 25.8 38.6





#5
Bromomethane
Concen: 51.936 ug/l
RT: 2.837 min Scan# 316
Delta R.T. -0.016 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 94 Resp: 326562
Ion Ratio Lower Upper
94 100
96 93.7 73.8 110.8



Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022






01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#6
Chloroethane
Concen: 50.088 ug/l
RT: 3.009 min Scan# 380
Delta R.T. -0.010 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 64 Resp: 367467
Ion Ratio Lower Upper
64 100
66 31.8 26.3 39.5





#7
 Trichlorofluoromethane
 Concen: 49.202 ug/l
 RT: 3.371 min Scan# 517
 Delta R.T. -0.005 min
 Lab File: VN070415.D
 Acq: 30 Dec 2021 22:48

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050





Tgt Ion:101 Resp: 72144
 Ion Ratio Lower Upper
 101 100
 103 64.5 50.1 75.1

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 01/03/2022
 Supervised By :mohammad ahmed 01/03/2022




01/03/2022
 Supervised By :mohammad
 ahmed


01/03/2022



#8
 Diethyl Ether
 Concen: 53.569 ug/l
 RT: 3.824 min Scan# 684
 Delta R.T. 0.000 min
 Lab File: VN070415.D
 Acq: 30 Dec 2021 22:48


Tgt Ion: 74 Resp: 346261
 Ion Ratio Lower Upper
 74 100
 45 113.8 47.0 141.2





#9
1,1,2-Trichlorotrifluoroethane
Concen: 52.395 ug/l
RT: 4.205 min Scan# 81
Delta R.T. -0.005 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion:101 Resp: 469894
Ion Ratio Lower Upper
101 100
85 46.2 37.2 55.8
151 77.3 65.9 98.9

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#10
Methyl Iodide
Concen: 54.760 ug/l
RT: 4.417 min Scan# 905
Delta R.T. -0.005 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion:142 Resp: 664620
Ion Ratio Lower Upper
142 100
127 43.3 35.4 53.0
141 14.3 11.6 17.4





#11
Tert butyl alcohol
Concen: 318.676 ug/l
RT: 5.388 min Scan# 11
Delta R.T. 0.011 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 59 Resp: 66172
Ion Ratio Lower Upper
59 100
57 10.2 8.7 13.1



Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022






01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#12
1,1-Dichloroethene
Concen: 52.848 ug/l
RT: 4.175 min Scan# 815
Delta R.T. -0.006 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 96 Resp: 466842
Ion Ratio Lower Upper
96 100
61 178.8 131.0 196.6
98 61.4 52.4 78.6





#13
Acrolein
Concen: 248.145 ug/l
RT: 4.039 min Scan# 764
Delta R.T. -0.002 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 418080
Ion Ratio Lower Upper
56 100
55 71.5 56.6 85.0


Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed



01/03/2022




#14
Allyl chloride
Concen: 56.791 ug/l
RT: 4.838 min Scan# 1062
Delta R.T. -0.002 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48




Tgt Ion: 41 Resp: 1137408
Ion Ratio Lower Upper
41 100
39 61.9 61.2 91.8
76 25.3 26.0 39.0#





#15
Acrylonitrile
Concen: 316.252 ug/l
RT: 5.557 min Scan# 11
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion: 53 Resp: 2049220
Ion Ratio Lower Upper
53 100
52 83.0 66.2 99.2
51 36.2 29.6 44.4

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#16
Acetone
Concen: 223.479 ug/l
RT: 4.291 min Scan# 858
Delta R.T. 0.006 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Tgt Ion: 43 Resp: 1809192
Ion Ratio Lower Upper
43 100
58 28.8 23.1 34.7








#17
Carbon Disulfide
Concen: 55.639 ug/l
RT: 4.524 min Scan# 947
Delta R.T. -0.005 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 76 Resp: 1341620
Ion Ratio Lower Upper
76 100
78 9.2 7.5 11.3


Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022






01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#18
Methyl Acetate
Concen: 61.168 ug/l
RT: 4.859 min Scan# 1070
Delta R.T. 0.005 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 43 Resp: 1432430
Ion Ratio Lower Upper
43 100
74 19.8 19.4 29.2





#19
Methyl tert-butyl Ether
Concen: 54.174 ug/l
RT: 5.618 min Scan# 1157
Delta R.T. 0.002 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion: 73 Resp: 1777189
Ion Ratio Lower Upper
73 100
57 23.9 16.8 25.2

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#20
Methylene Chloride
Concen: 53.009 ug/l
RT: 5.093 min Scan# 1157
Delta R.T. -0.002 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 84 Resp: 593746
Ion Ratio Lower Upper
84 100
49 142.2 93.5 140.3#
51 42.6 29.4 44.2
86 63.6 53.0 79.4





#21
trans-1,2-Dichloroethene
Concen: 53.968 ug/l
RT: 5.594 min Scan# 1345
Delta R.T. -0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion: 96 Resp: 50476
Ion Ratio Lower Upper
96 100
61 153.2 105.8 158.6
98 63.5 52.0 78.0

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#22
Diisopropyl ether
Concen: 56.910 ug/l
RT: 6.501 min Scan# 1682
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 45 Resp: 2220030
Ion Ratio Lower Upper
45 100
43 57.9 51.7 77.5
87 22.8 23.2 34.8#
59 10.7 9.4 14.0





#23
 Vinyl Acetate
 Concen: 290.516 ug/l
 RT: 6.436 min Scan# 1658
 Delta R.T. 0.002 min
 Lab File: VN070415.D
 Acq: 30 Dec 2021 22:48

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050






Tgt Ion: 43 Resp: 8792810
 Ion Ratio Lower Upper
 43 100
 86 8.4 8.5 12.7

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 01/03/2022
 Supervised By :mohammad ahmed 01/03/2022




01/03/2022
 Supervised By :mohammad
 ahmed


01/03/2022



#24
 1,1-Dichloroethane
 Concen: 54.677 ug/l
 RT: 6.388 min Scan# 1640
 Delta R.T. 0.000 min
 Lab File: VN070415.D
 Acq: 30 Dec 2021 22:48


Tgt Ion: 63 Resp: 1105944
 Ion Ratio Lower Upper
 63 100
 98 5.9 3.3 9.9
 100 3.7 2.1 6.5





#25
2-Butanone
Concen: 289.241 ug/l
RT: 7.340 min Scan# 1995
Delta R.T. 0.005 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 2948329
Ion Ratio Lower Upper
43 100
72 22.7 21.7 32.5


Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed



01/03/2022




#26
2,2-Dichloropropane
Concen: 44.940 ug/l
RT: 7.327 min Scan# 1990
Delta R.T. -0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48




Tgt Ion: 77 Resp: 717099
Ion Ratio Lower Upper
77 100
97 22.2 11.6 34.8





#27
cis-1,2-Dichloroethene
Concen: 53.972 ug/l
RT: 7.327 min Scan# 1989
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion: 96 Resp: 602339
Ion Ratio Lower Upper
96 100
61 160.1 0.0 275.0
98 64.3 0.0 129.6

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#28
Bromochloromethane
Concen: 51.570 ug/l
RT: 7.659 min Scan# 2114
Delta R.T. 0.002 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 49 Resp: 510128
Ion Ratio Lower Upper
49 100
129 1.7 0.0 4.0
130 61.8 67.8 101.8#





#29
Tetrahydrofuran
Concen: 321.234 ug/l
RT: 7.689 min Scan# 2125
Delta R.T. 0.006 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion: 42 Resp: 195364
Ion Ratio Lower Upper
42 100
72 36.6 36.3 54.5
71 34.7 34.7 52.1

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#30
Chloroform
Concen: 53.849 ug/l
RT: 7.820 min Scan# 2174
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 83 Resp: 1050779
Ion Ratio Lower Upper
83 100
85 64.8 51.9 77.9





#31
Cyclohexane
Concen: 52.368 ug/l
RT: 8.094 min Scan# 2276
Delta R.T. -0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion: 56 Resp: 103057
Ion Ratio Lower Upper
56 100
69 31.4 27.1 40.7
84 77.0 73.4 110.0

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#32
1,1,1-Trichloroethane
Concen: 53.738 ug/l
RT: 8.016 min Scan# 2247
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Tgt Ion: 97 Resp: 884154
Ion Ratio Lower Upper
97 100
99 64.4 50.1 75.1
61 49.8 35.2 52.8





#33

1,2-Dichloroethane-d4

Concen: 50.242 ug/l

RT: 8.440 min Scan# 2404

Delta R.T. 0.003 min

Lab File: VN070415.D

Acq: 30 Dec 2021 22:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 62656

Ion Ratio Lower Upper

65 100

67 52.6 0.0 103.0

Manual Integrations

APPROVED





Reviewed By :Amit Patel 01/03/2022

Supervised By :mohammad ahmed 01/03/2022

01/03/2022

Supervised By :mohammad
ahmed

01/03/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.968 min Scan# 2602

Delta R.T. 0.000 min

Lab File: VN070415.D

Acq: 30 Dec 2021 22:48




Tgt Ion:114 Resp: 1452023


Ion Ratio Lower Upper

114 100

63 23.2 0.0 37.8


88 16.3 0.0 27.4





#35
Dibromofluoromethane
Concen: 51.492 ug/l
RT: 8.021 min Scan# 21
Delta R.T. 0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion: 113 Resp: 45238
Ion Ratio Lower Upper
113 100
111 103.4 82.2 123.2
192 19.3 16.9 25.3

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#36
1,1-Dichloropropene
Concen: 53.020 ug/l
RT: 8.223 min Scan# 2324
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Tgt Ion: 75 Resp: 770570
Ion Ratio Lower Upper
75 100
110 32.8 18.4 55.2
77 30.7 24.4 36.6





#37
Ethyl Acetate
Concen: 60.178 ug/l
RT: 7.415 min Scan# 2022
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050





Tgt Ion: 43 Resp: 113366
Ion Ratio Lower Upper
43 100
61 12.6 11.5 17.3
70 8.6 7.9 11.9

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#38
Carbon Tetrachloride
Concen: 54.631 ug/l
RT: 8.206 min Scan# 2318
Delta R.T. 0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion:117 Resp: 758159
Ion Ratio Lower Upper
117 100
119 96.8 76.1 114.1
121 30.5 24.2 36.2





#39
Methylcyclohexane
Concen: 52.829 ug/l
RT: 9.462 min Scan# 21
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion: 83 Resp: 92953
Ion Ratio Lower Upper
83 100
55 84.4 58.1 87.1
98 43.2 38.4 57.6

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#40
Benzene
Concen: 54.355 ug/l
RT: 8.461 min Scan# 2413
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 78 Resp: 2421888
Ion Ratio Lower Upper
78 100
77 23.8 18.6 27.8





#41
Methacrylonitrile
Concen: 61.981 ug/l
RT: 7.641 min Scan# 2106
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion: 41 Resp: 584024
Ion Ratio Lower Upper
41 100
39 56.6 48.1 72.1
67 67.7 68.3 102.5
52 29.2 29.6 44.4

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#42
1,2-Dichloroethane
Concen: 52.298 ug/l
RT: 8.531 min Scan# 2439
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 62 Resp: 875862
Ion Ratio Lower Upper
62 100
98 8.3 0.0 17.6





#43
Isopropyl Acetate
Concen: 58.096 ug/l
RT: 8.563 min Scan# 2450
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion: 43 Resp: 1725529
Ion Ratio Lower Upper
43 100
61 23.2 20.4 30.6
87 10.9 11.2 16.8

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#44
Trichloroethene
Concen: 52.008 ug/l
RT: 9.218 min Scan# 2695
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Tgt Ion:130 Resp: 550605
Ion Ratio Lower Upper
130 100
95 103.6 0.0 191.2





#45

1,2-Dichloropropane

Concen: 55.564 ug/l

RT: 9.491 min Scan# 21

Delta R.T. 0.003 min

Lab File: VN070415.D

Acq: 30 Dec 2021 22:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 666580

Ion Ratio Lower Upper

63 100

65 31.5 24.7 37.1

Manual Integrations

APPROVED





Reviewed By :Amit Patel 01/03/2022

Supervised By :mohammad ahmed 01/03/2022

01/03/2022

Supervised By :mohammad
ahmed

01/03/2022



#46

Dibromomethane

Concen: 54.805 ug/l

RT: 9.580 min Scan# 2830

Delta R.T. 0.003 min

Lab File: VN070415.D

Acq: 30 Dec 2021 22:48




Tgt Ion: 93 Resp: 408834


Ion Ratio Lower Upper

93 100

95 83.1 65.9 98.9


174 97.5 84.6 127.0





#47
Bromodichloromethane
Concen: 58.248 ug/l
RT: 9.762 min Scan# 2898
Delta R.T. -0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion: 83 Resp: 862788
Ion Ratio Lower Upper
83 100
85 64.4 50.9 76.3
127 8.2 7.1 10.7

Manual Integrations
APPROVED


Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022

01/03/2022
Supervised By :mohammad
ahmed



01/03/2022




#48
Methyl methacrylate
Concen: 58.689 ug/l
RT: 9.561 min Scan# 2823
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48




Tgt Ion: 41 Resp: 831451
Ion Ratio Lower Upper
41 100
69 76.0 72.4 108.6
39 59.5 58.4 87.6





#49
1,4-Dioxane
Concen: 1327.563 ug/l
RT: 9.572 min Scan# 2827
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion: 88 Resp: 275312
Ion Ratio Lower Upper
88 100
43 34.5 23.4 35.2
58 77.2 50.6 75.8

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#50
Toluene-d8
Concen: 50.261 ug/l
RT: 10.443 min Scan# 3152
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 98 Resp: 1803849
Ion Ratio Lower Upper
98 100
100 64.3 50.4 75.6





#51
4-Methyl-2-Pentanone
Concen: 321.194 ug/l
RT: 10.333 min Scan# 3111
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 5994449
Ion Ratio Lower Upper
43 100
58 35.6 30.4 45.6


Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022






01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#52
Toluene
Concen: 55.138 ug/l
RT: 10.508 min Scan# 3176
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Tgt Ion: 92 Resp: 1487043
Ion Ratio Lower Upper
92 100
91 175.3 142.0 213.0







#53
t-1,3-Dichloropropene
Concen: 56.325 ug/l
RT: 10.719 min Scan# 3057
Delta R.T. 0.002 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 75 Resp: 903185
Ion Ratio Lower Upper
75 100
77 31.2 25.6 38.4



Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022






01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#54
cis-1,3-Dichloropropene
Concen: 56.602 ug/l
RT: 10.191 min Scan# 3058
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 75 Resp: 981844
Ion Ratio Lower Upper
75 100
77 31.0 26.4 39.6
39 57.9 42.5 63.7





#55
1,1,2-Trichloroethane
Concen: 55.922 ug/l
RT: 10.899 min Scan# 3321
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050





Tgt Ion: 97 Resp: 596738
Ion Ratio Lower Upper
97 100
83 90.7 67.8 101.6
85 58.0 45.4 68.2
99 62.8 49.4 74.0

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#56
Ethyl methacrylate
Concen: 59.562 ug/l
RT: 10.762 min Scan# 3271
Delta R.T. 0.002 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Tgt Ion: 69 Resp: 1050178
Ion Ratio Lower Upper
69 100
41 77.8 51.4 77.0#
39 41.8 34.6 52.0





#57

1,3-Dichloropropane

Concen: 55.881 ug/l

RT: 11.044 min Scan# 31

Delta R.T. -0.000 min

Lab File: VN070415.D

Acq: 30 Dec 2021 22:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 1054679

Ion Ratio Lower Upper

76 100





78 32.3 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Amit Patel 01/03/2022

Supervised By :mohammad ahmed 01/03/2022



#58

2-Chloroethyl Vinyl ether

Concen: 248.112 ug/l

RT: 10.041 min Scan# 3002

Delta R.T. -0.000 min

Lab File: VN070415.D




Acq: 30 Dec 2021 22:48


Tgt Ion: 63 Resp: 1116742

Ion Ratio Lower Upper

63 100


106 24.3 23.9 35.9





#59
2-Hexanone
Concen: 321.267 ug/l
RT: 11.087 min Scan# 3391
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 4372428
Ion Ratio Lower Upper
43 100
58 49.6 26.9 80.7


Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022






01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#60
Dibromochloromethane
Concen: 61.830 ug/l
RT: 11.240 min Scan# 3449
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion:129 Resp: 643732
Ion Ratio Lower Upper
129 100
127 77.0 38.6 116.0





#61
1,2-Dibromoethane
Concen: 56.906 ug/l
RT: 11.347 min Scan# 3488
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:107 Resp: 616574
Ion Ratio Lower Upper
107 100
109 93.5 75.7 113.5


Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022






01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#62
4-Bromofluorobenzene
Concen: 50.527 ug/l
RT: 12.731 min Scan# 4005
Delta R.T. -0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 95 Resp: 654952
Ion Ratio Lower Upper
95 100
174 76.9 0.0 187.2
176 73.3 0.0 181.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.747 min Scan# 3638
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion:117 Resp: 1395800
Ion Ratio Lower Upper
117 100
82 58.6 42.4 63.6
119 31.9 25.4 38.2

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#64
Tetrachloroethene
Concen: 48.539 ug/l
RT: 10.980 min Scan# 3352
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion:164 Resp: 490613
Ion Ratio Lower Upper
164 100
166 126.6 101.9 152.9
129 96.0 72.0 108.0
131 92.9 70.9 106.3





#65
Chlorobenzene
Concen: 52.266 ug/l
RT: 11.771 min Scan# 3646
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:112 Resp: 1547914
Ion Ratio Lower Upper
112 100
114 32.2 25.4 38.0



Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022






01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#66
1,1,1,2-Tetrachloroethane
Concen: 54.879 ug/l
RT: 11.841 min Scan# 3673
Delta R.T. 0.001 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion:131 Resp: 580221
Ion Ratio Lower Upper
131 100
133 95.1 47.9 143.7
119 65.1 32.1 96.5





#67
Ethyl Benzene
Concen: 53.707 ug/l
RT: 11.846 min Scan# 3674
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 91 Resp: 2914330
Ion Ratio Lower Upper
91 100
106 30.3 25.7 38.5


Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022






01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#68
m/p-Xylenes
Concen: 107.940 ug/l
RT: 11.953 min Scan# 3715
Delta R.T. 0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Tgt Ion:106 Resp: 2132038
Ion Ratio Lower Upper
106 100
91 210.4 160.6 240.8







#69
o-Xylene
Concen: 54.053 ug/l
RT: 12.280 min Scan# 3837
Delta R.T. 0.002 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 1061964
Ion Ratio Lower Upper
106 100
91 222.1 105.8 317.3



Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022






01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#70
Styrene
Concen: 56.796 ug/l
RT: 12.294 min Scan# 3842
Delta R.T. -0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion:104 Resp: 1766081
Ion Ratio Lower Upper
104 100
78 53.0 38.7 58.1
103 55.3 45.0 67.6





#71
Bromoform
Concen: 65.960 ug/l
RT: 12.457 min Scan# 3903
Delta R.T. 0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion:173 Resp: 51756
Ion Ratio Lower Upper
173 100
175 48.8 24.2 72.6
254 0.0 0.0 0.0

Manual Integrations
APPROVED


Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022

01/03/2022
Supervised By :mohammad
ahmed



01/03/2022




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.675 min Scan# 4357
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48




Tgt Ion:152 Resp: 550436
Ion Ratio Lower Upper
152 100
115 61.2 27.5 82.5
150 177.0 0.0 344.4





#73
Isopropylbenzene
Concen: 54.501 ug/l
RT: 12.578 min Scan# 3948
Delta R.T. 0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 105 Resp: 278589
Ion Ratio Lower Upper
105 100
120 25.7 13.4 40.1



Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022






01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#74
N-ethyl acetate
Concen: 60.140 ug/l
RT: 12.388 min Scan# 3877
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 43 Resp: 1292560
Ion Ratio Lower Upper
43 100
70 36.5 33.4 50.0
55 23.4 19.3 28.9
61 21.3 18.5 27.7





#75
1,1,2,2-Tetrachloroethane
Concen: 59.114 ug/l
RT: 12.827 min Scan# 4040
Delta R.T. 0.002 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion: 83 Resp: 97233
Ion Ratio Lower Upper
83 100
131 9.7 5.5 16.5
85 64.4 32.4 97.2

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#76
1,2,3-Trichloropropane
Concen: 57.800 ug/l m
RT: 12.881 min Scan# 4061
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 75 Resp: 949235
Ion Ratio Lower Upper
75 100
77 0.0 82.6 247.8#





#77
Bromobenzene
Concen: 54.274 ug/l
RT: 12.860 min Scan# 4053
Delta R.T. -0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion:156 Resp: 63692
Ion Ratio Lower Upper
156 100
77 233.2 90.6 271.8
158 97.2 48.9 146.7

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#78
n-propylbenzene
Concen: 55.652 ug/l
RT: 12.921 min Scan# 4076
Delta R.T. 0.002 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 91 Resp: 3151910
Ion Ratio Lower Upper
91 100
120 21.9 11.9 35.9





#79
2-Chlorotoluene
Concen: 53.250 ug/l
RT: 13.007 min Scan# 4108
Delta R.T. 0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 91 Resp: 1935069
Ion Ratio Lower Upper
91 100
126 32.0 18.3 54.8



Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022






01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#80
1,3,5-Trimethylbenzene
Concen: 54.506 ug/l
RT: 13.061 min Scan# 4128
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion:105 Resp: 2232675
Ion Ratio Lower Upper
105 100
120 47.5 24.6 73.6





#81
trans-1,4-Dichloro-2-butene
Concen: 64.845 ug/l
RT: 12.626 min Scan# 3965
Delta R.T. 0.002 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050





Tgt Ion: 75 Resp: 29214
Ion Ratio Lower Upper
75 100
53 108.0 90.5 135.7
89 44.5 38.2 57.4

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#82
4-Chlorotoluene
Concen: 53.808 ug/l
RT: 13.104 min Scan# 4144
Delta R.T. -0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 91 Resp: 1831594
Ion Ratio Lower Upper
91 100
126 32.1 18.2 54.6





#83
tert-Butylbenzene
Concen: 54.682 ug/l
RT: 13.324 min Scan# 41
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion:119 Resp: 197228
Ion Ratio Lower Upper
119 100
91 68.7 31.5 94.5
134 25.4 12.4 37.4

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#84
1,2,4-Trimethylbenzene
Concen: 55.763 ug/l
RT: 13.369 min Scan# 4243
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion:105 Resp: 2199486
Ion Ratio Lower Upper
105 100
120 43.8 22.8 68.4





#85
 sec-Butylbenzene
 Concen: 55.459 ug/l
 RT: 13.501 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VN070415.D
 Acq: 30 Dec 2021 22:48


Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050



Tgt Ion:105 Resp: 2715780
 Ion Ratio Lower Upper
 105 100
 134 19.2 10.4 31.1


Manual Integrations
 APPROVED

Reviewed By :Amit Patel 01/03/2022
 Supervised By :mohammad ahmed 01/03/2022






01/03/2022
 Supervised By :mohammad
 ahmed


01/03/2022



#86
 p-Isopropyltoluene
 Concen: 55.216 ug/l
 RT: 13.616 min Scan# 4335
 Delta R.T. 0.003 min
 Lab File: VN070415.D
 Acq: 30 Dec 2021 22:48


Tgt Ion:119 Resp: 2123989
 Ion Ratio Lower Upper
 119 100
 134 26.1 13.6 40.7
 91 24.6 11.4 34.1





#87
1,3-Dichlorobenzene
Concen: 53.391 ug/l
RT: 13.616 min Scan# 4334
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion:146 Resp: 1048880
Ion Ratio Lower Upper
146 100
111 41.4 18.6 55.6
148 63.4 31.6 95.0

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#88
1,4-Dichlorobenzene
Concen: 51.908 ug/l
RT: 13.694 min Scan# 4364
Delta R.T. -0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion:146 Resp: 1005907
Ion Ratio Lower Upper
146 100
111 39.7 18.3 54.8
148 63.8 31.8 95.4





#89
n-Butylbenzene
Concen: 53.656 ug/l
RT: 13.943 min Scan# 4456
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050





Tgt Ion: 91 Resp: 172156
Ion Ratio Lower Upper
91 100
92 52.5 25.4 76.0
134 24.5 13.6 40.8

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#90
Hexachloroethane
Concen: 67.167 ug/l
RT: 14.211 min Scan# 4557
Delta R.T. 0.002 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Tgt Ion:117 Resp: 448850
Ion Ratio Lower Upper
117 100
201 79.1 43.9 131.6





#91
1,2-Dichlorobenzene
Concen: 52.489 ug/l
RT: 13.989 min Scan# 4473
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050





Tgt Ion:146 Resp: 101088
Ion Ratio Lower Upper
146 100
111 42.5 19.1 57.3
148 64.1 32.1 96.3

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




01/03/2022
Supervised By :mohammad
ahmed


01/03/2022



#92
1,2-Dibromo-3-Chloropropane
Concen: 63.313 ug/l
RT: 14.603 min Scan# 4703
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48


Tgt Ion: 75 Resp: 187477
Ion Ratio Lower Upper
75 100
155 76.1 46.9 140.7
157 97.7 60.6 181.7





#93
1,2,4-Trichlorobenzene
Concen: 55.265 ug/l
RT: 15.263 min Scan# 4949
Delta R.T. -0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion:180 Resp: 50150
Ion Ratio Lower Upper
180 100
182 95.4 48.3 144.9
145 30.6 14.5 43.5

Manual Integrations
APPROVED


Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022

01/03/2022
Supervised By :mohammad
ahmed



01/03/2022




#94
Hexachlorobutadiene
Concen: 52.832 ug/l
RT: 15.370 min Scan# 4989
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48




Tgt Ion:225 Resp: 278635
Ion Ratio Lower Upper
225 100
223 62.8 31.9 95.7
227 63.2 32.1 96.3





#95
Naphthalene
Concen: 60.972 ug/l
RT: 15.507 min Scan# 5039
Delta R.T. 0.003 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050






Tgt Ion:128 Resp: 1539394
Ion Ratio Lower Upper
128 100
127 13.0 10.5 15.7
129 10.9 8.7 13.1

Manual Integrations
APPROVED

Reviewed By :Amit Patel 01/03/2022
Supervised By :mohammad ahmed 01/03/2022

01/03/2022
Supervised By :mohammad
ahmed

01/03/2022



#96
1,2,3-Trichlorobenzene
Concen: 58.619 ug/l
RT: 15.700 min Scan# 5112
Delta R.T. -0.000 min
Lab File: VN070415.D
Acq: 30 Dec 2021 22:48

Tgt Ion:180 Resp: 510073
Ion Ratio Lower Upper
180 100
182 94.3 47.4 142.2
145 32.1 15.0 45.1

