



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123021\
Data File : VN070414.D
Acq On : 30 Dec 2021 21:31
Operator : JC/MD
Sample : BFB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Title : SW846 8260
Last Update : Mon Dec 27 18:47:12 2021



AutoFind: Scans 4005, 4006, 4007; Background Corrected with Scan 3987

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.1	46419	PASS
75	95	30	60	53.5	117659	PASS
95	95	100	100	100.0	219840	PASS
96	95	5	9	6.3	13757	PASS
173	174	0.00	2	0.9	1530	PASS
174	95	50	100	76.5	168213	PASS
175	174	5	9	7.6	12732	PASS
176	174	95	101	97.9	164651	PASS
177	176	5	9	6.3	10429	PASS