

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123024\
 Data File : VN085354.D
 Acq On : 30 Dec 2024 18:51
 Operator : JC\MD
 Sample : P5386-01ME 10X
 Misc : 4.99g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MOO-24-00398ME

Quant Time: Dec 31 02:18:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 28 01:36:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	127049	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	238668	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	208676	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	100424	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	110918	61.210	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 122.420%			
35) Dibromofluoromethane	8.165	113	83378	55.998	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 112.000%			
50) Toluene-d8	10.565	98	291201	53.297	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.600%			
62) 4-Bromofluorobenzene	12.847	95	95870	52.834	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 105.660%			
Target Compounds						
16) Acetone	4.442	43	21000	35.243	ug/l	92
52) Toluene	10.629	92	7174	1.828	ug/l	91
67) Ethyl Benzene	11.971	91	3843	0.512	ug/l	96
68) m/p-Xylenes	12.065	106	9256	3.373	ug/l	98
69) o-Xylene	12.400	106	18524	6.977	ug/l	92
73) Isopropylbenzene	12.688	105	11683	1.618	ug/l	99
78) n-propylbenzene	13.035	91	70214	8.341	ug/l	100
80) 1,3,5-Trimethylbenzene	13.170	105	95498	16.349	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	378855	65.136	ug/l	98
85) sec-Butylbenzene	13.617	105	12379	1.762	ug/l #	50
86) p-Isopropyltoluene	13.729	119	72680	12.253	ug/l	95
89) n-Butylbenzene	14.053	91	9836	1.912	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123024\
Data File : VN085354.D
Acq On : 30 Dec 2024 18:51
Operator : JC\MD
Sample : P5386-01ME 10X
Misc : 4.99g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MOO-24-00398ME

Quant Time: Dec 31 02:18:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
Quant Title : SW846 8260
QLast Update : Sat Dec 28 01:36:16 2024
Response via : Initial Calibration

