

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123024\
 Data File : VN085356.D
 Acq On : 30 Dec 2024 19:39
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 31 02:19:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 28 01:36:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	144292	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	269542	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	237668	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	98721	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	121594	59.083	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	118.160%
35) Dibromofluoromethane	8.165	113	93462	55.581	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.160%
50) Toluene-d8	10.565	98	333399	54.031	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.060%
62) 4-Bromofluorobenzene	12.847	95	103672	50.589	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123024\
Data File : VN085356.D
Acq On : 30 Dec 2024 19:39
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 31 02:19:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
Quant Title : SW846 8260
QLast Update : Sat Dec 28 01:36:16 2024
Response via : Initial Calibration

