

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123024\
 Data File : VN085341.D
 Acq On : 30 Dec 2024 13:26
 Operator : JC\MD
 Sample : VN1230WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1230WBL01

Quant Time: Dec 31 02:11:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 28 01:36:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	161031	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	303169	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	251914	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	97891	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	131504	57.256	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	114.520%	
35) Dibromofluoromethane	8.165	113	99877	52.808	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	105.620%	
50) Toluene-d8	10.565	98	354787	51.119	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.240%	
62) 4-Bromofluorobenzene	12.847	95	107174	46.497	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	93.000%	


Target Compounds Qvalue


 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123024\
Data File : VN085341.D
Acq On : 30 Dec 2024 13:26
Operator : JC\MD
Sample : VN1230WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1230WBL01


Quant Time: Dec 31 02:11:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
Quant Title : SW846 8260
QLast Update : Sat Dec 28 01:36:16 2024
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085341.D
Acq: 30 Dec 2024 13:26

Instrument :
MSVOA_N
ClientSampleId :
VN1230WBL01







Tgt Ion:168 Resp: 161031
Ion Ratio Lower Upper
168 100
99 66.0 51.0 76.4



#33
1,2-Dichloroethane-d4
Concen: 57.256 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. 0.000 min
Lab File: VN085341.D
Acq: 30 Dec 2024 13:26

Tgt Ion: 65 Resp: 131504
Ion Ratio Lower Upper
65 100
67 50.2 0.0 104.0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 9.100 min Scan# 1214
 Delta R.T. 0.006 min
 Lab File: VN085341.D
 Acq: 30 Dec 2024 13:26




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1230WBL01


Tgt Ion:	114	Resp:	303169
Ion	Ratio	Lower	Upper
114	100		
63	22.7	0.0	43.4
88	14.5	0.0	33.6



#35
 Dibromofluoromethane
 Concen: 52.808 ug/l
 RT: 8.165 min Scan# 1056
 Delta R.T. 0.006 min
 Lab File: VN085341.D
 Acq: 30 Dec 2024 13:26

Tgt Ion:	113	Resp:	99877
Ion	Ratio	Lower	Upper
113	100		
111	102.8	80.5	120.7
192	17.0	12.9	19.3





#50

Toluene-d8

Concen: 51.119 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.006 min

Lab File: VN085341.D

Acq: 30 Dec 2024 13:26

Instrument :

MSVOA_N

ClientSampleId :





VN1230WBL01

Tgt Ion: 98 Resp: 354787

Ion Ratio Lower Upper

98 100

100 63.1 51.5 77.3



#62

4-Bromofluorobenzene

Concen: 46.497 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. 0.000 min

Lab File: VN085341.D

Acq: 30 Dec 2024 13:26




Tgt Ion: 95 Resp: 107174


Ion Ratio Lower Upper

95 100

174 72.1 0.0 143.2

176 70.3 0.0 135.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1684
Delta R.T. 0.006 min
Lab File: VN085341.D
Acq: 30 Dec 2024 13:26

Instrument :
MSVOA_N
ClientSampleId :
VN1230WBL01

Tgt Ion:117 Resp: 251914
Ion Ratio Lower Upper
117 100
82 56.9 45.7 68.5
119 31.4 25.8 38.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN085341.D
Acq: 30 Dec 2024 13:26

Tgt Ion:152 Resp: 97891
Ion Ratio Lower Upper
152 100
115 62.5 30.4 91.3
150 158.0 0.0 345.0

