

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
 Data File : VN065343.D
 Acq On : 31 Dec 2020 14:13
 Operator : JC/MD
 Sample : VN1231WBL02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1231WBL02

Quant Time: Jan 01 01:49:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	258044	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	433045	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	455310	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	176080	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	149266	51.24	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.48%	
35) Dibromofluoromethane	7.553	113	132608	50.76	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.52%	
50) Toluene-d8	10.058	98	535371	54.67	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 109.34%	
62) 4-Bromofluorobenzene	12.376	95	203311	52.25	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 104.50%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
Data File : VN065343.D
Acq On : 31 Dec 2020 14:13
Operator : JC/MD
Sample : VN1231WBL02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1231WBL02

Quant Time: Jan 01 01:49:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

