

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
 Data File : VN065346.D
 Acq On : 31 Dec 2020 15:30
 Operator : JC/MD
 Sample : PB133875ZHE#20
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133875ZHE#20

Quant Time: Jan 01 01:50:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	267550	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	445655	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	480434	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	192389	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.994	65	150640	49.87	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.74%
35) Dibromofluoromethane	7.556	113	136195	50.66	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.32%
50) Toluene-d8	10.061	98	555845	55.16	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	110.32%
62) 4-Bromofluorobenzene	12.376	95	215855	53.74	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.48%
Target Compounds						
20) Methylene Chloride	4.515	84	5580	1.89	ug/l	96
43) Isopropyl Acetate	8.138	43	25824	3.73	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
Data File : VN065346.D
Acq On : 31 Dec 2020 15:30
Operator : JC/MD
Sample : PB133875ZHE#20
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133875ZHE#20

Quant Time: Jan 01 01:50:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

