

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
 Data File : VN065349.D
 Acq On : 31 Dec 2020 16:44
 Operator : JC/MD
 Sample : PB133875ZHE#23
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133875ZHE#23

Quant Time: Jan 01 01:51:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.631	168	255071	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	431941	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	464388	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	186885	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.994	65	148399	51.54	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 103.08%			
35) Dibromofluoromethane	7.557	113	130361	50.02	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 100.04%			
50) Toluene-d8	10.061	98	540276	55.32	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 110.64%			
62) 4-Bromofluorobenzene	12.376	95	207329	53.30	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 106.60%			
Target Compounds						
20) Methylene Chloride	4.512	84	5674	2.02	ug/l	96
43) Isopropyl Acetate	8.135	43	26065	3.88	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
Data File : VN065349.D
Acq On : 31 Dec 2020 16:44
Operator : JC/MD
Sample : PB133875ZHE#23
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133875ZHE#23

Quant Time: Jan 01 01:51:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

