

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
 Data File : VN065350.D
 Acq On : 31 Dec 2020 17:09
 Operator : JC/MD
 Sample : PB133875ZHE#24
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133875ZHE#24

Quant Time: Jan 01 01:51:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.631	168	260085	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	436314	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	473922	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.319	152	191127	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.994	65	149156	50.80	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 101.60%	
35) Dibromofluoromethane	7.557	113	132478	50.33	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.66%	
50) Toluene-d8	10.061	98	550786	55.83	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 111.66%	
62) 4-Bromofluorobenzene	12.376	95	214652	54.50	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 109.00%	
Target Compounds						
20) Methylene Chloride	4.512	84	6148	2.15	ug/l	Qvalue # 88
43) Isopropyl Acetate	8.135	43	26770	3.95	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
Data File : VN065350.D
Acq On : 31 Dec 2020 17:09
Operator : JC/MD
Sample : PB133875ZHE#24
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133875ZHE#24

Quant Time: Jan 01 01:51:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

