

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
 Data File : VN065352.D
 Acq On : 31 Dec 2020 17:59
 Operator : JC/MD
 Sample : L5241-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JG-OWS-4

Quant Time: Jan 01 01:52:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	257210	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	436982	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	471324	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	216105	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	148834	51.26	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.52%	
35) Dibromofluoromethane	7.553	113	133071	50.48	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.96%	
50) Toluene-d8	10.061	98	546496	55.31	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 110.62%	
62) 4-Bromofluorobenzene	12.376	95	219980	55.65	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 111.30%	
Target Compounds						
16) Acetone	3.791	43	46125	28.31	ug/l	95
20) Methylene Chloride	4.512	84	13780	4.86	ug/l	97
25) 2-Butanone	6.814	43	10001	6.14	ug/l	95
43) Isopropyl Acetate	8.138	43	16325	2.41	ug/l	94
52) Toluene	10.126	92	519600	69.75	ug/l	100
80) 1,3,5-Trimethylbenzene	12.707	105	16018	1.14	ug/l	99
84) 1,2,4-Trimethylbenzene	13.013	105	53317	3.90	ug/l	99
86) p-Isopropyltoluene	13.264	119	425342	29.25	ug/l	97
95) Naphthalene	15.103	128	46262	7.22	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
Data File : VN065352.D
Acq On : 31 Dec 2020 17:59
Operator : JC/MD
Sample : L5241-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JG-OWS-4

Quant Time: Jan 01 01:52:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

