

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
 Data File : VN065359.D
 Acq On : 31 Dec 2020 20:53
 Operator : JC/MD
 Sample : L5242-16
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SAMPLE-2(484+26)

Quant Time: Jan 01 01:54:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	259503	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	430843	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	466214	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	191768	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.994	65	148837	50.80	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 101.60%			
35) Dibromofluoromethane	7.556	113	132634	51.03	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 102.06%			
50) Toluene-d8	10.061	98	545388	55.98	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 111.96%			
62) 4-Bromofluorobenzene	12.376	95	209501	53.93	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 107.86%			
Target Compounds						
20) Methylene Chloride	4.515	84	11721	4.10	ug/l	97
43) Isopropyl Acetate	8.135	43	22053	3.30	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
Data File : VN065359.D
Acq On : 31 Dec 2020 20:53
Operator : JC/MD
Sample : L5242-16
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SAMPLE-2(484+26)

Quant Time: Jan 01 01:54:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

