

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
 Data File : VN065360.D
 Acq On : 31 Dec 2020 21:18
 Operator : JC/MD
 Sample : L5242-17
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SAMPLE-3(474+26)

Quant Time: Jan 01 01:55:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	256485	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	424258	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	457238	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	191579	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.994	65	146330	50.54	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.08%
35) Dibromofluoromethane	7.557	113	129967	50.78	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.56%
50) Toluene-d8	10.061	98	535366	55.81	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	111.62%
62) 4-Bromofluorobenzene	12.376	95	207848	54.30	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.60%
Target Compounds						
16) Acetone	3.798	43	21571	7.31	ug/l	97
20) Methylene Chloride	4.515	84	13029	4.61	ug/l	99
43) Isopropyl Acetate	8.135	43	20204	3.07	ug/l	95
95) Naphthalene	15.106	128	32005	6.38	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
Data File : VN065360.D
Acq On : 31 Dec 2020 21:18
Operator : JC/MD
Sample : L5242-17
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SAMPLE-3(474+26)

Quant Time: Jan 01 01:55:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

