

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
 Data File : VN065361.D
 Acq On : 31 Dec 2020 21:42
 Operator : JC/MD
 Sample : L5242-18
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SAMPLE-4(464+26)

Quant Time: Jan 01 01:55:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	259968	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	439100	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	472819	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	195210	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.994	65	149748	51.02	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.04%	
35) Dibromofluoromethane	7.553	113	133216	50.29	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.58%	
50) Toluene-d8	10.061	98	543872	54.78	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 109.56%	
62) 4-Bromofluorobenzene	12.376	95	214263	54.10	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 108.20%	
Target Compounds						
						Qvalue
20) Methylene Chloride	4.518	84	12671	4.43	ug/l	94
43) Isopropyl Acetate	8.135	43	22620	3.32	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
Data File : VN065361.D
Acq On : 31 Dec 2020 21:42
Operator : JC/MD
Sample : L5242-18
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SAMPLE-4(464+26)

Quant Time: Jan 01 01:55:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

