

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
 Data File : VN065362.D
 Acq On : 31 Dec 2020 22:07
 Operator : JC/MD
 Sample : L5242-19
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SAMPLE-5(454+26)

Quant Time: Jan 01 01:55:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	258755	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	437009	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	470939	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	196356	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.994	65	149537	51.19	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 102.38%			
35) Dibromofluoromethane	7.556	113	132553	50.28	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 100.56%			
50) Toluene-d8	10.061	98	547588	55.42	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 110.84%			
62) 4-Bromofluorobenzene	12.376	95	212846	54.01	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 108.02%			
Target Compounds						
20) Methylene Chloride	4.508	84	11416	4.01	ug/l	98
43) Isopropyl Acetate	8.135	43	20236	2.98	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
Data File : VN065362.D
Acq On : 31 Dec 2020 22:07
Operator : JC/MD
Sample : L5242-19
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SAMPLE-5(454+26)

Quant Time: Jan 01 01:55:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

