

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
 Data File : VN065363.D
 Acq On : 31 Dec 2020 22:32
 Operator : JC/MD
 Sample : L5242-20
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SAMPLE-6(444+26)

Quant Time: Jan 01 01:56:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.631	168	258296	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	433429	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	464466	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	189618	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.994	65	148793	51.03	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 102.06%			
35) Dibromofluoromethane	7.557	113	131835	50.42	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 100.84%			
50) Toluene-d8	10.061	98	542307	55.33	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 110.66%			
62) 4-Bromofluorobenzene	12.376	95	209250	53.58	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 107.16%			
Target Compounds						Qvalue
18) Methyl Acetate	4.300	43	2842	1.09	ug/l #	68
20) Methylene Chloride	4.515	84	12737	4.48	ug/l	94
43) Isopropyl Acetate	8.135	43	15990	2.38	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
Data File : VN065363.D
Acq On : 31 Dec 2020 22:32
Operator : JC/MD
Sample : L5242-20
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SAMPLE-6(444+26)

Quant Time: Jan 01 01:56:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

