

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
 Data File : VN065364.D
 Acq On : 31 Dec 2020 22:57
 Operator : JC/MD
 Sample : L5242-21
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SAMPLE-7(434+26)

Quant Time: Jan 01 01:56:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.631	168	257467	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	436395	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	469782	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	195718	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.994	65	150029	51.62	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.24%	
35) Dibromofluoromethane	7.557	113	131947	50.12	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.24%	
50) Toluene-d8	10.061	98	544578	55.19	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 110.38%	
62) 4-Bromofluorobenzene	12.376	95	211002	53.65	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 107.30%	
Target Compounds						
						Qvalue
18) Methyl Acetate	4.303	43	3081	1.18	ug/l	# 72
20) Methylene Chloride	4.512	84	12947	4.57	ug/l	95
43) Isopropyl Acetate	8.139	43	19560	2.89	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
Data File : VN065364.D
Acq On : 31 Dec 2020 22:57
Operator : JC/MD
Sample : L5242-21
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SAMPLE-7(434+26)

Quant Time: Jan 01 01:56:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

