

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
 Data File : VN065342.D
 Acq On : 31 Dec 2020 13:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 01 01:48:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	118	0.00
2 T	Dichlorodifluoromethane	0.463	0.421	9.1	100	0.00
3 P	Chloromethane	0.527	0.520	1.3	111	0.00
4 C	Vinyl Chloride	0.548	0.543	0.9#	111	0.00
5 T	Bromomethane	0.389	0.379	2.6	112	0.00
6 T	Chloroethane	0.354	0.351	0.8	111	0.00
7 T	Trichlorofluoromethane	0.820	0.728	11.2	101	0.00
8 T	Diethyl Ether	0.323	0.291	9.9	107	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.460	0.423	8.0	107	0.00
10 T	Methyl Iodide	0.694	0.647	6.8	104	0.00
11 T	Tert butyl alcohol	0.092	0.084	8.7	103	0.00
12 CM	1,1-Dichloroethene	0.491	0.439	10.6#	106	0.00
13 T	Acrolein	0.065	0.055	15.4	101	0.00
14 T	Allyl chloride	0.810	0.701	13.5	99	0.00
15 T	Acrylonitrile	0.223	0.224	-0.4	112	0.00
16 T	Acetone	0.260	0.180	30.8#	87	0.00
17 T	Carbon Disulfide	1.257	1.100	12.5	98	0.00
18 T	Methyl Acetate	0.507	0.493	2.8	111	0.00
19 T	Methyl tert-butyl Ether	1.560	1.422	8.8	105	0.00
20 T	Methylene Chloride	0.551	0.497	9.8	113	0.00
21 T	trans-1,2-Dichloroethene	0.482	0.453	6.0	107	0.00
22 T	Diisopropyl ether	1.611	1.503	6.7	109	0.00
23 T	Vinyl Acetate	1.358	1.253	7.7	104	0.00
24 P	1,1-Dichloroethane	0.884	0.849	4.0	109	0.00
25 T	2-Butanone	0.317	0.292	7.9	104	0.00
26 T	2,2-Dichloropropane	0.844	0.680	19.4	95	0.00
27 T	cis-1,2-Dichloroethene	0.561	0.556	0.9	114	0.00
28 T	Bromochloromethane	0.379	0.351	7.4	109	0.00
29 T	Tetrahydrofuran	0.205	0.193	5.9	109	0.00
30 C	Chloroform	0.910	0.876	3.7#	111	0.00
31 T	Cyclohexane	0.833	0.732	12.1	106	0.00
32 T	1,1,1-Trichloroethane	0.824	0.756	8.3	105	0.00
33 S	1,2-Dichloroethane-d4	0.564	0.521	7.6	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	115	0.00
35 S	Dibromofluoromethane	0.302	0.303	-0.3	114	0.00
36 T	1,1-Dichloropropene	0.455	0.439	3.5	110	0.00
37 T	Ethyl Acetate	0.451	0.420	6.9	104	0.00
38 T	Carbon Tetrachloride	0.477	0.449	5.9	107	0.00
39 T	Methylcyclohexane	0.546	0.525	3.8	107	0.00
40 TM	Benzene	1.352	1.357	-0.4	114	0.00
41 T	Methacrylonitrile	0.234	0.213	9.0	108	0.00
42 TM	1,2-Dichloroethane	0.500	0.460	8.0	106	0.00
43 T	Isopropyl Acetate	0.777	0.709	8.8	102	0.00
44 TM	Trichloroethene	0.390	0.390	0.0	113	0.00
45 C	1,2-Dichloropropane	0.352	0.356	-1.1#	114	0.00
46 T	Dibromomethane	0.239	0.236	1.3	113	0.00
47 T	Bromodichloromethane	0.503	0.482	4.2	108	0.00
48 T	Methyl methacrylate	0.377	0.351	6.9	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
 Data File : VN065342.D
 Acq On : 31 Dec 2020 13:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 01 01:48:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.007	-16.7	121	0.00
50 S	Toluene-d8	1.131	1.152	-1.9	116	0.00
51 T	4-Methyl-2-Pentanone	0.437	0.427	2.3	109	0.00
52 CM	Toluene	0.852	0.868	-1.9#	115	0.00
53 T	t-1,3-Dichloropropene	0.567	0.535	5.6	106	0.00
54 T	cis-1,3-Dichloropropene	0.603	0.575	4.6	109	0.00
55 T	1,1,2-Trichloroethane	0.337	0.346	-2.7	116	0.00
56 T	Ethyl methacrylate	0.522	0.517	1.0	111	0.00
57 T	1,3-Dichloropropane	0.557	0.574	-3.1	115	0.00
58 T	2-Chloroethyl Vinyl ether	0.227	0.218	4.0	110	0.00
59 T	2-Hexanone	0.321	0.313	2.5	107	0.00
60 T	Dibromochloromethane	0.402	0.397	1.2	111	0.00
61 T	1,2-Dibromoethane	0.362	0.368	-1.7	114	0.00
62 S	4-Bromofluorobenzene	0.424	0.423	0.2	115	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	117	0.00
64 T	Tetrachloroethene	0.486	0.464	4.5	114	0.00
65 PM	Chlorobenzene	1.018	1.030	-1.2	117	0.00
66 T	1,1,1,2-Tetrachloroethane	0.396	0.395	0.3	114	0.00
67 C	Ethyl Benzene	1.834	1.820	0.8#	115	0.00
68 T	m/p-Xylenes	0.691	0.699	-1.2	117	0.00
69 T	o-Xylene	0.671	0.674	-0.4	116	0.00
70 T	Styrene	1.119	1.151	-2.9	117	0.00
71 P	Bromoform	0.341	0.334	2.1	111	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	118	0.00
73 T	Isopropylbenzene	3.808	3.603	5.4	116	0.00
74 T	N-amyl acetate	1.516	1.370	9.6	105	0.00
75 P	1,1,2,2-Tetrachloroethane	1.027	1.029	-0.2	121	0.00
76 T	1,2,3-Trichloropropane	1.020	1.006	1.4	125	0.00
77 T	Bromobenzene	0.995	0.979	1.6	119	0.00
78 T	n-propylbenzene	4.299	4.049	5.8	114	0.00
79 T	2-Chlorotoluene	2.532	2.402	5.1	114	0.00
80 T	1,3,5-Trimethylbenzene	3.247	3.088	4.9	116	0.00
81 T	trans-1,4-Dichloro-2-butene	0.405	0.359	11.4	104	0.00
82 T	4-Chlorotoluene	2.607	2.514	3.6	116	0.00
83 T	tert-Butylbenzene	2.767	2.629	5.0	115	0.00
84 T	1,2,4-Trimethylbenzene	3.162	3.090	2.3	117	0.00
85 T	sec-Butylbenzene	3.671	3.481	5.2	115	0.00
86 T	p-Isopropyltoluene	3.365	3.194	5.1	114	0.00
87 T	1,3-Dichlorobenzene	1.664	1.684	-1.2	120	0.00
88 T	1,4-Dichlorobenzene	1.671	1.650	1.3	117	0.00
89 T	n-Butylbenzene	2.876	2.649	7.9	108	0.00
90 T	Hexachloroethane	0.579	0.553	4.5	111	0.00
91 T	1,2-Dichlorobenzene	1.598	1.646	-3.0	120	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.212	0.205	3.3	110	0.00
93 T	1,2,4-Trichlorobenzene	0.806	0.920	-14.1	115	0.00
94 T	Hexachlorobutadiene	0.631	0.565	10.5	112	0.00
95 T	Naphthalene	2.105	2.349	-11.6	115	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123120\
Data File : VN065342.D
Acq On : 31 Dec 2020 13:05
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 01 01:48:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.764	0.862	-12.8	119	0.02

(#) = Out of Range SPCC's out = 0 CCC's out = 6