

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123124\
 Data File : VN085362.D
 Acq On : 31 Dec 2024 10:19
 Operator : JC\MD
 Sample : VN1231WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1231WBL01

Quant Time: Jan 01 02:01:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 28 01:36:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	139418	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	261209	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	224661	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	78057	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	119764	60.228	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 120.460%		
35) Dibromofluoromethane	8.165	113	94311	57.875	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 115.760%		
50) Toluene-d8	10.565	98	320302	53.564	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 107.120%		
62) 4-Bromofluorobenzene	12.847	95	89221	44.926	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 89.860%		

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN123124\
Data File : VN085362.D
Acq On : 31 Dec 2024 10:19
Operator : JC\MD
Sample : VN1231WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1231WBL01

Quant Time: Jan 01 02:01:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
Quant Title : SW846 8260
QLast Update : Sat Dec 28 01:36:16 2024
Response via : Initial Calibration

