

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
 Data File : VN077097.D
 Acq On : 22 Mar 2023 10:40
 Operator : JC\MD
 Sample : VN0322MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0322MBL01

Quant Time: Mar 22 23:19:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	342546	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	630503	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	547598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	202247	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	276746	50.579	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.160%
35) Dibromofluoromethane	8.172	113	202365	45.831	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.660%
50) Toluene-d8	10.571	98	755620	47.907	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.820%
62) 4-Bromofluorobenzene	12.854	95	249029	44.681	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
Data File : VN077097.D
Acq On : 22 Mar 2023 10:40
Operator : JC\MD
Sample : VN0322MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0322MBL01

Quant Time: Mar 22 23:19:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

