

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
 Data File : VN077104.D
 Acq On : 22 Mar 2023 13:41
 Operator : JC\MD
 Sample : 01849-01
 Misc : 12.53 gr/5.0mL/100uL/5.0mL/MSVOA_N/MeOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CB-10-9.0-030823

Quant Time: Mar 22 23:23:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.225	168	322478	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.101	114	584739	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	509574	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	185981	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.578	65	261756	50.816	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.640%
35) Dibromofluoromethane	8.166	113	185540	45.309	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.620%
50) Toluene-d8	10.566	98	717645	49.060	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.120%
62) 4-Bromofluorobenzene	12.854	95	227351	43.984	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.960%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.337	101	20849	4.872	ug/l	100
27) cis-1,2-Dichloroethene	7.489	96	66678	13.719	ug/l	92
44) Trichloroethene	9.354	130	1311684	294.392	ug/l	94
64) Tetrachloroethene	11.107	164	10352	3.027	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
Data File : VN077104.D
Acq On : 22 Mar 2023 13:41
Operator : JC\MD
Sample : 01849-01
Misc : 12.53 gr/5.0mL/100uL/5.0mL/MSVOA_N/MeOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CB-10-9.0-030823

Quant Time: Mar 22 23:23:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

