

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
 Data File : VN077106.D
 Acq On : 22 Mar 2023 14:29
 Operator : JC\MD
 Sample : 01849-03
 Misc : 11.50 gr/5.0mL/100uL/5.0mL/MSVOA_N/MeOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CB-10-17.5-030823

Quant Time: Mar 22 23:24:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.225	168	345465	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.101	114	626383	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	559335	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	212838	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.578	65	276831	50.167	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.340%	
35) Dibromofluoromethane	8.172	113	196112	44.707	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 89.420%	
50) Toluene-d8	10.572	98	779660	49.756	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.520%	
62) 4-Bromofluorobenzene	12.854	95	252023	45.515	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.040%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	4.301	96	4490	1.071	ug/l	# 88
27) cis-1,2-Dichloroethene	7.484	96	13171	2.530	ug/l	97
44) Trichloroethene	9.354	130	2324114	486.940	ug/l	92
64) Tetrachloroethene	11.107	164	5595	1.490	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
Data File : VN077106.D
Acq On : 22 Mar 2023 14:29
Operator : JC\MD
Sample : 01849-03
Misc : 11.50 gr/5.0mL/100uL/5.0mL/MSVOA_N/MeOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CB-10-17.5-030823

Quant Time: Mar 22 23:24:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

