

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
 Data File : VN077107.D
 Acq On : 22 Mar 2023 14:53
 Operator : JC\MD
 Sample : 01849-04
 Misc : 10.42 gr/5.0mL/100uL/5.0mL/MSVOA_N/MeOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CB-10-22.0-030823

Quant Time: Mar 22 23:24:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.225	168	340975	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.101	114	630067	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	569594	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	223993	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	275236	50.535	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.060%
35) Dibromofluoromethane	8.166	113	195089	44.213	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.420%
50) Toluene-d8	10.572	98	773513	49.075	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.160%
62) 4-Bromofluorobenzene	12.854	95	261112	46.881	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.760%
Target Compounds						
44) Trichloroethene	9.354	130	387043	80.618	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
Data File : VN077107.D
Acq On : 22 Mar 2023 14:53
Operator : JC\MD
Sample : 01849-04
Misc : 10.42 gr/5.0mL/100uL/5.0mL/MSVOA_N/MeOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CB-10-22.0-030823

Quant Time: Mar 22 23:24:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

