

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
 Data File : VN077122.D
 Acq On : 23 Mar 2023 02:39
 Operator : JC\MD
 Sample : VN0322WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0322WBL02

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 03/23/2023
 Supervised By :Semsettin Yesilyurt 03/23/2023

Quant Time: Mar 23 03:44:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	426360	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	722636	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	622179	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	239474	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	304602	44.726	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.460%
35) Dibromofluoromethane	8.172	113	234582	46.354	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.700%
50) Toluene-d8	10.571	98	864375	47.815	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.620%
62) 4-Bromofluorobenzene	12.854	95	262071m	41.026	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.060%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
Data File : VN077122.D
Acq On : 23 Mar 2023 02:39
Operator : JC\MD
Sample : VN0322WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0322WBL02

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 03/23/2023
Supervised By :Semsettin Yesilyurt 03/23/2023

Quant Time: Mar 23 03:44:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

