

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
 Data File : VN077127.D
 Acq On : 23 Mar 2023 04:59
 Operator : JC\MD
 Sample : PB151586ZHE003
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151586ZHE003

Quant Time: Mar 23 09:29:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	375856	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	662545	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	600472	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	222622	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	275356	45.865	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	91.720%		
35) Dibromofluoromethane	8.172	113	209227	45.093	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.180%		
50) Toluene-d8	10.572	98	833075	50.263	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	100.520%		
62) 4-Bromofluorobenzene	12.854	95	253597	43.300	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	86.600%		
Target Compounds						
16) Acetone	4.448	43	40892	15.286	ug/l	97
20) Methylene Chloride	5.295	84	17647	3.040	ug/l	93
37) Ethyl Acetate	7.578	43	80986	10.445	ug/l	98
43) Isopropyl Acetate	8.695	43	378112	31.216	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
Data File : VN077127.D
Acq On : 23 Mar 2023 04:59
Operator : JC\MD
Sample : PB151586ZHE003
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151586ZHE003

Quant Time: Mar 23 09:29:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

