

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
 Data File : VN077128.D
 Acq On : 23 Mar 2023 05:23
 Operator : JC\MD
 Sample : PB151586ZHE004
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151586ZHE004

Quant Time: Mar 23 09:29:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	406517	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	741213	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	669180	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	243268	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	316448	48.734	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.460%
35) Dibromofluoromethane	8.172	113	235158	45.303	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.600%
50) Toluene-d8	10.572	98	932658	50.299	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.600%
62) 4-Bromofluorobenzene	12.854	95	276720	42.233	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.460%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	39630	13.697	ug/l #	88
20) Methylene Chloride	5.290	84	19214	3.060	ug/l	92
37) Ethyl Acetate	7.572	43	77316	8.913	ug/l	96
43) Isopropyl Acetate	8.695	43	361537	26.680	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
Data File : VN077128.D
Acq On : 23 Mar 2023 05:23
Operator : JC\MD
Sample : PB151586ZHE004
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151586ZHE004

Quant Time: Mar 23 09:29:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

