

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
 Data File : VN077132.D
 Acq On : 23 Mar 2023 07:00
 Operator : JC\MD
 Sample : PB151586ZHE008
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151586ZHE008

Quant Time: Mar 23 09:30:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	384796	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	724713	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	668298	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	262736	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	306203	49.818	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.640%		
35) Dibromofluoromethane	8.172	113	230234	45.364	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.720%		
50) Toluene-d8	10.572	98	915572	50.502	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	101.000%		
62) 4-Bromofluorobenzene	12.854	95	280379	43.766	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	87.540%		
Target Compounds					Qvalue	
16) Acetone	4.454	43	30844	11.262	ug/l	91
20) Methylene Chloride	5.289	84	14160	2.382	ug/l	90
37) Ethyl Acetate	7.572	43	75096	8.854	ug/l	99
43) Isopropyl Acetate	8.695	43	402213	30.357	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
Data File : VN077132.D
Acq On : 23 Mar 2023 07:00
Operator : JC\MD
Sample : PB151586ZHE008
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151586ZHE008

Quant Time: Mar 23 09:30:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

