

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
 Data File : VN077133.D
 Acq On : 23 Mar 2023 07:24
 Operator : JC\MD
 Sample : PB151586ZHE009
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151586ZHE009

Quant Time: Mar 23 09:30:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	411270	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	759988	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	691799	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	255456	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	319634	48.656	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.320%
35) Dibromofluoromethane	8.172	113	242810	45.621	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.240%
50) Toluene-d8	10.571	98	947611	49.843	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.680%
62) 4-Bromofluorobenzene	12.854	95	287378	42.776	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.560%
Target Compounds						
					Qvalue	
16) Acetone	4.454	43	30220	10.324	ug/l	91
20) Methylene Chloride	5.301	84	15680	2.468	ug/l	94
37) Ethyl Acetate	7.578	43	74886	8.420	ug/l	97
43) Isopropyl Acetate	8.695	43	405865	29.211	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
Data File : VN077133.D
Acq On : 23 Mar 2023 07:24
Operator : JC\MD
Sample : PB151586ZHE009
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151586ZHE009

Quant Time: Mar 23 09:30:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

