

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
 Data File : VN077096.D
 Acq On : 22 Mar 2023 10:02
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 22 23:18:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.566	0.343	39.4#	71	0.00
3 P	Chloromethane	0.504	0.335	33.5#	70	0.00
4 C	Vinyl Chloride	0.544	0.458	15.8#	87	0.00
5 T	Bromomethane	0.367	0.343	6.5	92	0.00
6 T	Chloroethane	0.344	0.297	13.7	94	0.00
7 T	Trichlorofluoromethane	1.170	1.106	5.5	97	0.00
8 T	Diethyl Ether	0.440	0.421	4.3	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.664	0.611	8.0	97	0.00
10 T	Methyl Iodide	0.747	0.708	5.2	92	0.00
11 T	Tert butyl alcohol	0.116	0.096	17.2	81	0.00
12 CM	1,1-Dichloroethene	0.607	0.565	6.9#	94	0.00
13 T	Acrolein	0.144	0.133	7.6	93	0.00
14 T	Allyl chloride	1.111	1.180	-6.2	104	0.00
15 T	Acrylonitrile	0.391	0.349	10.7	88	0.00
16 T	Acetone	0.356	0.371	-4.2	108	0.00
17 T	Carbon Disulfide	1.770	1.520	14.1	89	0.00
18 T	Methyl Acetate	1.342	1.194	11.0	90	0.00
19 T	Methyl tert-butyl Ether	2.179	2.142	1.7	97	0.00
20 T	Methylene Chloride	0.772	0.695	10.0	97	0.00
21 T	trans-1,2-Dichloroethene	0.658	0.612	7.0	94	0.00
22 T	Diisopropyl ether	2.435	2.466	-1.3	98	0.00
23 T	Vinyl Acetate	1.536	1.536	0.0	95	0.00
24 P	1,1-Dichloroethane	1.393	1.318	5.4	97	0.00
25 T	2-Butanone	0.521	0.482	7.5	90	0.00
26 T	2,2-Dichloropropane	1.048	1.070	-2.1	100	0.00
27 T	cis-1,2-Dichloroethene	0.754	0.751	0.4	98	0.00
28 T	Bromochloromethane	0.605	0.545	9.9	96	0.00
29 T	Tetrahydrofuran	0.337	0.308	8.6	87	0.00
30 C	Chloroform	1.352	1.290	4.6#	98	0.00
31 T	Cyclohexane	1.307	1.195	8.6	95	0.00
32 T	1,1,1-Trichloroethane	1.156	1.116	3.5	97	0.00
33 S	1,2-Dichloroethane-d4	0.799	0.792	0.9	100	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.350	0.337	3.7	98	0.00
36 T	1,1-Dichloropropene	0.547	0.543	0.7	95	0.00
37 T	Ethyl Acetate	0.585	0.555	5.1	93	0.00
38 T	Carbon Tetrachloride	0.533	0.527	1.1	96	0.00
39 T	Methylcyclohexane	0.640	0.669	-4.5	95	0.00
40 TM	Benzene	1.659	1.617	2.5	96	0.00
41 T	Methacrylonitrile	0.315	0.321	-1.9	100	0.00
42 TM	1,2-Dichloroethane	0.599	0.581	3.0	97	0.00
43 T	Isopropyl Acetate	0.914	0.905	1.0	92	0.00
44 TM	Trichloroethene	0.381	0.357	6.3	94	0.00
45 C	1,2-Dichloropropane	0.441	0.441	0.0	98	0.00
46 T	Dibromomethane	0.278	0.268	3.6	96	0.00
47 T	Bromodichloromethane	0.583	0.584	-0.2	99	0.00
48 T	Methyl methacrylate	0.437	0.452	-3.4	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
 Data File : VN077096.D
 Acq On : 22 Mar 2023 10:02
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 22 23:18:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.006	14.3	81	0.00
50 S	Toluene-d8	1.251	1.286	-2.8	98	0.00
51 T	4-Methyl-2-Pentanone	0.569	0.547	3.9	89	0.00
52 CM	Toluene	0.990	0.994	-0.4#	95	0.00
53 T	t-1,3-Dichloropropene	0.571	0.617	-8.1	96	0.00
54 T	cis-1,3-Dichloropropene	0.644	0.682	-5.9	98	0.00
55 T	1,1,2-Trichloroethane	0.391	0.387	1.0	97	0.00
56 T	Ethyl methacrylate	0.587	0.622	-6.0	93	0.00
57 T	1,3-Dichloropropane	0.715	0.700	2.1	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.250	0.233	6.8	83	0.00
59 T	2-Hexanone	0.413	0.405	1.9	90	0.00
60 T	Dibromochloromethane	0.392	0.396	-1.0	94	0.00
61 T	1,2-Dibromoethane	0.387	0.378	2.3	95	0.00
62 S	4-Bromofluorobenzene	0.442	0.483	-9.3	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.336	0.307	8.6	92	0.00
65 PM	Chlorobenzene	1.150	1.111	3.4	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.426	0.412	3.3	97	0.00
67 C	Ethyl Benzene	2.090	2.164	-3.5#	97	0.00
68 T	m/p-Xylenes	0.775	0.803	-3.6	97	0.00
69 T	o-Xylene	0.770	0.788	-2.3	96	0.00
70 T	Styrene	1.199	1.297	-8.2	96	0.00
71 P	Bromoform	0.288	0.286	0.7	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	4.481	4.427	1.2	96	0.00
74 T	N-amyl acetate	1.939	1.872	3.5	92	0.00
75 P	1,1,2,2-Tetrachloroethane	1.623	1.358	16.3	93	0.00
76 T	1,2,3-Trichloropropane	1.293	1.040	19.6	79	0.00
77 T	Bromobenzene	1.004	0.931	7.3	98	0.00
78 T	n-propylbenzene	5.173	5.359	-3.6	98	0.00
79 T	2-Chlorotoluene	3.296	3.209	2.6	97	0.00
80 T	1,3,5-Trimethylbenzene	3.741	3.796	-1.5	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.470	0.435	7.4	92	0.00
82 T	4-Chlorotoluene	3.134	3.168	-1.1	100	0.00
83 T	tert-Butylbenzene	3.141	3.155	-0.4	96	0.00
84 T	1,2,4-Trimethylbenzene	3.733	3.817	-2.3	98	0.00
85 T	sec-Butylbenzene	4.468	4.697	-5.1	97	0.00
86 T	p-Isopropyltoluene	3.588	3.812	-6.2	98	0.00
87 T	1,3-Dichlorobenzene	1.897	1.795	5.4	99	0.00
88 T	1,4-Dichlorobenzene	1.887	1.760	6.7	97	0.00
89 T	n-Butylbenzene	3.148	3.506	-11.4	100	0.00
90 T	Hexachloroethane	0.730	0.693	5.1	99	0.00
91 T	1,2-Dichlorobenzene	1.891	1.735	8.2	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.301	0.255	15.3	83	0.00
93 T	1,2,4-Trichlorobenzene	0.855	0.898	-5.0	93	0.00
94 T	Hexachlorobutadiene	0.471	0.450	4.5	96	0.00
95 T	Naphthalene	2.729	2.856	-4.7	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032223\
Data File : VN077096.D
Acq On : 22 Mar 2023 10:02
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 22 23:18:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.858	0.858	0.0	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5