

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032423\
 Data File : VN077170.D
 Acq On : 24 Mar 2023 00:59
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: Mar 24 05:57:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032423W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Mar 24 05:55:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.825	128	40955	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.107	114	265376	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.872	117	281976	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.584	65	129556	31.149	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.833%
60) 4-Bromofluorobenzene	12.854	95	118379	23.369	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	77.900%
63) Toluene-d8	10.572	98	298749	29.680	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.933%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032423\
Data File : VN077170.D
Acq On : 24 Mar 2023 00:59
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: Mar 24 05:57:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032423W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Mar 24 05:55:38 2023
Response via : Initial Calibration

