

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033019\
 Data File : VN054796.D
 Acq On : 30 Mar 2019 12:27
 Operator : JC/SP
 Sample : VN0330WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0330WBL01

Quant Time: Mar 30 22:51:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	538389	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	856242	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	729086	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	256736	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	352616	48.59	ug/l	0.00
Spiked Amount						
			Recovery	=	97.18%	
35) Dibromofluoromethane	7.59	113	272630	47.76	ug/l	0.00
Spiked Amount						
			Recovery	=	95.52%	
50) Toluene-d8	10.09	98	1041297	49.06	ug/l	0.00
Spiked Amount						
			Recovery	=	98.12%	
62) 4-Bromofluorobenzene	12.40	95	321997	45.62	ug/l	0.00
Spiked Amount						
			Recovery	=	91.24%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN033019\
Data File : VN054796.D
Acq On : 30 Mar 2019 12:27
Operator : JC/SP
Sample : VN0330WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0330WBL01

Quant Time: Mar 30 22:51:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

