

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033019\
 Data File : VN054805.D
 Acq On : 30 Mar 2019 16:25
 Operator : JC/SP
 Sample : K2181-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4

Quant Time: Mar 30 23:23:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	517297	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	862520	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	692457	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	239023	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	349641	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
35) Dibromofluoromethane	7.59	113	273010	47.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.96%	
50) Toluene-d8	10.09	98	1008160	47.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.32%	
62) 4-Bromofluorobenzene	12.40	95	296456	41.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.40%	
Target Compounds						
11) Tert butyl alcohol	4.84	59	22796	62.743	ug/l #	79
19) Methyl tert-butyl Ether	5.06	73	33062	2.324	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN033019\
Data File : VN054805.D
Acq On : 30 Mar 2019 16:25
Operator : JC/SP
Sample : K2181-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4

Quant Time: Mar 30 23:23:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

