

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033019\
 Data File : VN054809.D
 Acq On : 30 Mar 2019 18:12
 Operator : JC/SP
 Sample : K2138-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S6B(25)

Quant Time: Mar 30 23:46:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	541566	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	862666	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	739044	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	273860	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	362463	49.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.30%	
35) Dibromofluoromethane	7.59	113	281400	48.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.86%	
50) Toluene-d8	10.09	98	1042844	48.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.54%	
62) 4-Bromofluorobenzene	12.40	95	319285	44.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.80%	
Target Compounds						
16) Acetone	3.82	43	8968	5.151	ug/l #	81
20) Methylene Chloride	4.54	84	20918	3.406	ug/l	89
43) Isopropyl Acetate	8.17	43	16504	1.723	ug/l #	95
95) Naphthalene	15.13	128	90786	9.198	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN033019\
Data File : VN054809.D
Acq On : 30 Mar 2019 18:12
Operator : JC/SP
Sample : K2138-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S6B(25)

Quant Time: Mar 30 23:46:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

