

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033019\
 Data File : VN054811.D
 Acq On : 30 Mar 2019 19:05
 Operator : JC/SP
 Sample : K2138-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S6A(10)

Quant Time: Mar 30 23:50:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	533004	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	885323	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	745433	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	267860	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	368255	51.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.50%	
35) Dibromofluoromethane	7.59	113	286176	48.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.96%	
50) Toluene-d8	10.09	98	1053199	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
62) 4-Bromofluorobenzene	12.40	95	321483	44.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.10%	
Target Compounds						
16) Acetone	3.82	43	8772	5.120	ug/l #	77
20) Methylene Chloride	4.55	84	23515	3.891	ug/l #	91
43) Isopropyl Acetate	8.18	43	17571	1.788	ug/l #	93
95) Naphthalene	15.13	128	9656	1.000	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN033019\
Data File : VN054811.D
Acq On : 30 Mar 2019 19:05
Operator : JC/SP
Sample : K2138-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S6A(10)

Quant Time: Mar 30 23:50:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

