

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033019\
 Data File : VN054814.D
 Acq On : 30 Mar 2019 20:25
 Operator : JC/SP
 Sample : K2138-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S6B(8)

Quant Time: Mar 30 23:57:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	537926	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	881239	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	741971	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	274777	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	365611	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
35) Dibromofluoromethane	7.59	113	278500	47.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.80%	
50) Toluene-d8	10.09	98	1046392	47.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.82%	
62) 4-Bromofluorobenzene	12.40	95	324245	44.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.28%	
Target Compounds						
20) Methylene Chloride	4.55	84	20166	3.306	ug/l #	88
43) Isopropyl Acetate	8.17	43	16202	1.656	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN033019\
Data File : VN054814.D
Acq On : 30 Mar 2019 20:25
Operator : JC/SP
Sample : K2138-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S6B(8)

Quant Time: Mar 30 23:57:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

