

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033019\
 Data File : VN054820.D
 Acq On : 30 Mar 2019 23:58
 Operator : JC/SP
 Sample : VN0330WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0330WBL02

Quant Time: Mar 31 04:55:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	527706	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	863404	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	713039	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	252270	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	356825	50.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.32%	
35) Dibromofluoromethane	7.59	113	274232	47.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.28%	
50) Toluene-d8	10.09	98	1031229	48.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.38%	
62) 4-Bromofluorobenzene	12.40	95	308320	43.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.64%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN033019\
Data File : VN054820.D
Acq On : 30 Mar 2019 23:58
Operator : JC/SP
Sample : VN0330WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0330WBL02

Quant Time: Mar 31 04:55:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

