

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033019\
 Data File : VN054829.D
 Acq On : 31 Mar 2019 3:57
 Operator : JC/SP
 Sample : K2164-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-6

Quant Time: Mar 31 05:11:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	520116	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	852500	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	711477	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	242310	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	352620	50.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.58%	
35) Dibromofluoromethane	7.59	113	274763	48.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.68%	
50) Toluene-d8	10.09	98	1014863	48.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.06%	
62) 4-Bromofluorobenzene	12.40	95	298332	42.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.90%	
Target Compounds						
11) Tert butyl alcohol	4.83	59	10725	29.359	ug/l #	70
19) Methyl tert-butyl Ether	5.06	73	262723	18.371	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN033019\
Data File : VN054829.D
Acq On : 31 Mar 2019 3:57
Operator : JC/SP
Sample : K2164-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-6

Quant Time: Mar 31 05:11:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

