

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033019\
 Data File : VN054831.D
 Acq On : 31 Mar 2019 4:50
 Operator : JC/SP
 Sample : K2153-28
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP-2

Quant Time: Apr 01 01:54:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	538691	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	877361	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	741191	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	266977	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	362063	49.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.72%	
35) Dibromofluoromethane	7.59	113	280040	47.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.76%	
50) Toluene-d8	10.09	98	1056677	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
62) 4-Bromofluorobenzene	12.40	95	320646	44.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.68%	
Target Compounds						
20) Methylene Chloride	4.55	84	29065	4.758	ug/l	86
43) Isopropyl Acetate	8.18	43	14623	1.501	ug/l	98
95) Naphthalene	15.13	128	16343	1.698	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN033019\
Data File : VN054831.D
Acq On : 31 Mar 2019 4:50
Operator : JC/SP
Sample : K2153-28
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP-2

Quant Time: Apr 01 01:54:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

