

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033019\
 Data File : VN054833.D
 Acq On : 31 Mar 2019 5:43
 Operator : JC/SP
 Sample : K2153-24
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP-1

Quant Time: Apr 01 02:02:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	511995	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	856724	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	717497	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	256682	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	354807	51.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.82%	
35) Dibromofluoromethane	7.59	113	275592	48.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.50%	
50) Toluene-d8	10.09	98	1018393	47.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.92%	
62) 4-Bromofluorobenzene	12.40	95	305735	43.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.58%	
Target Compounds						
20) Methylene Chloride	4.55	84	89591	15.432	ug/l #	84
43) Isopropyl Acetate	8.17	43	14491	1.523	ug/l	96
95) Naphthalene	15.13	128	22649	2.448	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN033019\
Data File : VN054833.D
Acq On : 31 Mar 2019 5:43
Operator : JC/SP
Sample : K2153-24
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP-1

Quant Time: Apr 01 02:02:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

