

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033019\
 Data File : VN054834.D
 Acq On : 31 Mar 2019 6:10
 Operator : JC/SP
 Sample : K1685-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AU-06-032819-C

Quant Time: Apr 01 02:04:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	526766	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	876930	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	744042	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	269641	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	361369	50.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.78%	
35) Dibromofluoromethane	7.59	113	283412	48.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.96%	
50) Toluene-d8	10.09	98	1042411	47.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.92%	
62) 4-Bromofluorobenzene	12.40	95	322906	44.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.34%	
Target Compounds						
16) Acetone	3.83	43	11314	6.682	ug/l #	82
20) Methylene Chloride	4.55	84	168067	28.138	ug/l #	80
43) Isopropyl Acetate	8.17	43	15321	1.574	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN033019\
Data File : VN054834.D
Acq On : 31 Mar 2019 6:10
Operator : JC/SP
Sample : K1685-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AU-06-032819-C

Quant Time: Apr 01 02:04:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

