

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033019\
 Data File : VN054835.D
 Acq On : 31 Mar 2019 6:37
 Operator : JC/SP
 Sample : K2139-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S7B(8)

Quant Time: Apr 01 02:07:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	534966	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	885322	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	757459	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	270232	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	361736	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
35) Dibromofluoromethane	7.59	113	281807	47.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.48%	
50) Toluene-d8	10.09	98	1058402	48.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.46%	
62) 4-Bromofluorobenzene	12.40	95	322911	44.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.50%	
Target Compounds						
20) Methylene Chloride	4.54	84	19877	3.277	ug/l #	86
43) Isopropyl Acetate	8.17	43	16627	1.692	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN033019\
Data File : VN054835.D
Acq On : 31 Mar 2019 6:37
Operator : JC/SP
Sample : K2139-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S7B(8)

Quant Time: Apr 01 02:07:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

