

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061722\
 Data File : VN072919.D
 Acq On : 17 Jun 2022 19:56
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: Jun 18 06:24:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	299711	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	530148	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	445113	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	182634	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	211587	52.218	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.440%
35) Dibromofluoromethane	7.951	113	162044	49.452	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.900%
50) Toluene-d8	10.375	98	596409	48.914	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.820%
62) 4-Bromofluorobenzene	12.669	95	198424	45.256	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061722\
Data File : VN072919.D
Acq On : 17 Jun 2022 19:56
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: Jun 18 06:24:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

