

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061722\  
 Data File : VN072918.D  
 Acq On : 17 Jun 2022 19:32  
 Operator : JC\MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 ICSVN061722

Quant Time: Jun 18 06:23:44 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N061722W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 18 06:22:45 2022  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|       | Compound                    | Amount  | Calc.   | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|---------|---------|-------|-------|----------|
| 1 I   | Pentafluorobenzene          | 50.000  | 50.000  | 0.0   | 100   | 0.00     |
| 2 T   | Dichlorodifluoromethane     | 50.000  | 52.522  | -5.0  | 100   | 0.00     |
| 3 P   | Chloromethane               | 50.000  | 50.704  | -1.4  | 104   | 0.00     |
| 4 C   | Vinyl Chloride              | 50.000  | 51.120  | -2.2# | 99    | 0.00     |
| 5 T   | Bromomethane                | 50.000  | 58.929  | -17.9 | 107   | 0.00     |
| 6 T   | Chloroethane                | 50.000  | 50.070  | -0.1  | 100   | 0.00     |
| 7 T   | Trichlorofluoromethane      | 50.000  | 45.756  | 8.5   | 98    | 0.00     |
| 8 T   | Diethyl Ether               | 50.000  | 50.438  | -0.9  | 99    | 0.00     |
| 9 T   | 1,1,2-Trichlorotrifluoroeth | 50.000  | 48.523  | 3.0   | 97    | 0.00     |
| 10 T  | Methyl Iodide               | 50.000  | 52.487  | -5.0  | 104   | 0.00     |
| 11 T  | Tert butyl alcohol          | 250.000 | 261.901 | -4.8  | 102   | 0.00     |
| 12 CM | 1,1-Dichloroethene          | 50.000  | 50.121  | -0.2# | 99    | 0.00     |
| 13 T  | Acrolein                    | 250.000 | 255.982 | -2.4  | 102   | 0.00     |
| 14 T  | Allyl chloride              | 50.000  | 49.797  | 0.4   | 100   | 0.00     |
| 15 T  | Acrylonitrile               | 250.000 | 260.372 | -4.1  | 102   | 0.00     |
| 16 T  | Acetone                     | 250.000 | 254.124 | -1.6  | 104   | 0.00     |
| 17 T  | Carbon Disulfide            | 50.000  | 50.138  | -0.3  | 101   | 0.00     |
| 18 T  | Methyl Acetate              | 50.000  | 48.281  | 3.4   | 101   | 0.00     |
| 19 T  | Methyl tert-butyl Ether     | 50.000  | 51.960  | -3.9  | 100   | 0.00     |
| 20 T  | Methylene Chloride          | 50.000  | 50.180  | -0.4  | 98    | 0.00     |
| 21 T  | trans-1,2-Dichloroethene    | 50.000  | 48.866  | 2.3   | 99    | 0.00     |
| 22 T  | Diisopropyl ether           | 50.000  | 51.850  | -3.7  | 100   | 0.00     |
| 23 T  | Vinyl Acetate               | 250.000 | 264.244 | -5.7  | 101   | 0.00     |
| 24 P  | 1,1-Dichloroethane          | 50.000  | 50.950  | -1.9  | 100   | 0.00     |
| 25 T  | 2-Butanone                  | 250.000 | 259.843 | -3.9  | 102   | 0.00     |
| 26 T  | 2,2-Dichloropropane         | 50.000  | 46.322  | 7.4   | 88    | 0.00     |
| 27 T  | cis-1,2-Dichloroethene      | 50.000  | 50.130  | -0.3  | 101   | 0.00     |
| 28 T  | Bromochloromethane          | 50.000  | 53.504  | -7.0  | 102   | 0.00     |
| 29 T  | Tetrahydrofuran             | 250.000 | 263.251 | -5.3  | 104   | 0.00     |
| 30 C  | Chloroform                  | 50.000  | 50.447  | -0.9# | 99    | 0.00     |
| 31 T  | Cyclohexane                 | 50.000  | 47.137  | 5.7   | 100   | 0.00     |
| 32 T  | 1,1,1-Trichloroethane       | 50.000  | 51.410  | -2.8  | 100   | 0.00     |
| 33 S  | 1,2-Dichloroethane-d4       | 50.000  | 51.984  | -4.0  | 104   | 0.00     |
| 34 I  | 1,4-Difluorobenzene         | 50.000  | 50.000  | 0.0   | 102   | 0.00     |
| 35 S  | Dibromofluoromethane        | 50.000  | 52.036  | -4.1  | 104   | 0.00     |
| 36 T  | 1,1-Dichloropropene         | 50.000  | 49.652  | 0.7   | 100   | 0.00     |
| 37 T  | Ethyl Acetate               | 50.000  | 50.541  | -1.1  | 103   | 0.00     |
| 38 T  | Carbon Tetrachloride        | 50.000  | 49.786  | 0.4   | 100   | 0.00     |
| 39 T  | Methylcyclohexane           | 50.000  | 49.186  | 1.6   | 97    | 0.00     |
| 40 TM | Benzene                     | 50.000  | 50.163  | -0.3  | 100   | 0.00     |
| 41 T  | Methacrylonitrile           | 50.000  | 50.305  | -0.6  | 104   | 0.00     |
| 42 TM | 1,2-Dichloroethane          | 50.000  | 51.416  | -2.8  | 101   | 0.00     |
| 43 T  | Isopropyl Acetate           | 50.000  | 51.819  | -3.6  | 101   | 0.00     |
| 44 TM | Trichloroethene             | 50.000  | 48.963  | 2.1   | 100   | 0.00     |
| 45 C  | 1,2-Dichloropropane         | 50.000  | 50.577  | -1.2# | 101   | 0.00     |
| 46 T  | Dibromomethane              | 50.000  | 51.144  | -2.3  | 101   | 0.00     |
| 47 T  | Bromodichloromethane        | 50.000  | 52.163  | -4.3  | 102   | 0.00     |
| 48 T  | Methyl methacrylate         | 50.000  | 50.954  | -1.9  | 101   | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061722\  
 Data File : VN072918.D  
 Acq On : 17 Jun 2022 19:32  
 Operator : JC\MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 ICSVN061722

Quant Time: Jun 18 06:23:44 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N061722W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 18 06:22:45 2022  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|       | Compound                    | Amount   | Calc.    | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|----------|----------|-------|-------|----------|
| 49 T  | 1,4-Dioxane                 | 1000.000 | 1014.656 | -1.5  | 102   | 0.00     |
| 50 S  | Toluene-d8                  | 50.000   | 52.951   | -5.9  | 105   | 0.00     |
| 51 T  | 4-Methyl-2-Pentanone        | 250.000  | 261.537  | -4.6  | 102   | 0.00     |
| 52 CM | Toluene                     | 50.000   | 51.058   | -2.1# | 101   | 0.00     |
| 53 T  | t-1,3-Dichloropropene       | 50.000   | 52.873   | -5.7  | 99    | 0.00     |
| 54 T  | cis-1,3-Dichloropropene     | 50.000   | 51.779   | -3.6  | 100   | 0.00     |
| 55 T  | 1,1,2-Trichloroethane       | 50.000   | 48.360   | 3.3   | 101   | 0.00     |
| 56 T  | Ethyl methacrylate          | 50.000   | 53.329   | -6.7  | 102   | 0.00     |
| 57 T  | 1,3-Dichloropropane         | 50.000   | 49.892   | 0.2   | 101   | 0.00     |
| 58 T  | 2-Chloroethyl Vinyl ether   | 250.000  | 271.342  | -8.5  | 100   | 0.00     |
| 59 T  | 2-Hexanone                  | 250.000  | 266.608  | -6.6  | 104   | 0.00     |
| 60 T  | Dibromochloromethane        | 50.000   | 54.381   | -8.8  | 101   | 0.00     |
| 61 T  | 1,2-Dibromoethane           | 50.000   | 52.812   | -5.6  | 103   | 0.00     |
| 62 S  | 4-Bromofluorobenzene        | 50.000   | 53.003   | -6.0  | 105   | 0.00     |
| 63 I  | Chlorobenzene-d5            | 50.000   | 50.000   | 0.0   | 101   | 0.00     |
| 64 T  | Tetrachloroethene           | 50.000   | 49.079   | 1.8   | 100   | 0.00     |
| 65 PM | Chlorobenzene               | 50.000   | 50.511   | -1.0  | 100   | 0.00     |
| 66 T  | 1,1,1,2-Tetrachloroethane   | 50.000   | 51.091   | -2.2  | 101   | 0.00     |
| 67 C  | Ethyl Benzene               | 50.000   | 52.166   | -4.3# | 101   | 0.00     |
| 68 T  | m/p-Xylenes                 | 100.000  | 101.798  | -1.8  | 100   | 0.00     |
| 69 T  | o-Xylene                    | 50.000   | 51.191   | -2.4  | 101   | 0.00     |
| 70 T  | Styrene                     | 50.000   | 52.366   | -4.7  | 101   | 0.00     |
| 71 P  | Bromoform                   | 50.000   | 54.723   | -9.4  | 101   | 0.00     |
| 72 I  | 1,4-Dichlorobenzene-d4      | 50.000   | 50.000   | 0.0   | 101   | 0.00     |
| 73 T  | Isopropylbenzene            | 50.000   | 49.894   | 0.2   | 101   | 0.00     |
| 74 T  | N-ethyl acetate             | 50.000   | 51.304   | -2.6  | 102   | 0.00     |
| 75 P  | 1,1,2,2-Tetrachloroethane   | 50.000   | 49.095   | 1.8   | 101   | 0.00     |
| 76 T  | 1,2,3-Trichloropropane      | 50.000   | 44.531   | 10.9  | 101   | 0.00     |
| 77 T  | Bromobenzene                | 50.000   | 49.576   | 0.8   | 99    | 0.00     |
| 78 T  | n-propylbenzene             | 50.000   | 51.973   | -3.9  | 101   | 0.00     |
| 79 T  | 2-Chlorotoluene             | 50.000   | 49.542   | 0.9   | 101   | 0.00     |
| 80 T  | 1,3,5-Trimethylbenzene      | 50.000   | 51.208   | -2.4  | 101   | 0.00     |
| 81 T  | trans-1,4-Dichloro-2-butene | 50.000   | 51.236   | -2.5  | 98    | 0.00     |
| 82 T  | 4-Chlorotoluene             | 50.000   | 50.762   | -1.5  | 101   | 0.00     |
| 83 T  | tert-Butylbenzene           | 50.000   | 50.533   | -1.1  | 101   | 0.00     |
| 84 T  | 1,2,4-Trimethylbenzene      | 50.000   | 50.761   | -1.5  | 101   | 0.00     |
| 85 T  | sec-Butylbenzene            | 50.000   | 51.355   | -2.7  | 101   | 0.00     |
| 86 T  | p-Isopropyltoluene          | 50.000   | 51.712   | -3.4  | 99    | 0.00     |
| 87 T  | 1,3-Dichlorobenzene         | 50.000   | 49.496   | 1.0   | 101   | 0.00     |
| 88 T  | 1,4-Dichlorobenzene         | 50.000   | 48.479   | 3.0   | 99    | 0.00     |
| 89 T  | n-Butylbenzene              | 50.000   | 53.021   | -6.0  | 100   | 0.00     |
| 90 T  | Hexachloroethane            | 50.000   | 52.182   | -4.4  | 102   | 0.00     |
| 91 T  | 1,2-Dichlorobenzene         | 50.000   | 48.977   | 2.0   | 101   | 0.00     |
| 92 T  | 1,2-Dibromo-3-Chloropropane | 50.000   | 50.957   | -1.9  | 102   | 0.00     |
| 93 T  | 1,2,4-Trichlorobenzene      | 50.000   | 51.426   | -2.9  | 102   | 0.00     |
| 94 T  | Hexachlorobutadiene         | 50.000   | 49.632   | 0.7   | 100   | 0.00     |
| 95 T  | Naphthalene                 | 50.000   | 50.449   | -0.9  | 106   | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061722\  
Data File : VN072918.D  
Acq On : 17 Jun 2022 19:32  
Operator : JC\MD  
Sample : VSTDICV050  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
ICVVN061722

Quant Time: Jun 18 06:23:44 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N061722W.M  
Quant Title : SW846 8260  
QLast Update : Sat Jun 18 06:22:45 2022  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound               | Amount | Calc.  | %Dev | Area% | Dev(min) |
|------|------------------------|--------|--------|------|-------|----------|
| 96 T | 1,2,3-Trichlorobenzene | 50.000 | 50.581 | -1.2 | 106   | 0.00     |

(#) = Out of Range                      SPCC's out = 0    CCC's out = 6