

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082622\
 Data File : VN074132.D
 Acq On : 26 Aug 2022 13:12
 Operator : JC\MD
 Sample : VN0826WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0826WBL01

Quant Time: Aug 27 00:19:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 06:44:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	630296	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	943758	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	792145	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	341729	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	322972	52.609	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.220%
35) Dibromofluoromethane	7.951	113	281593	50.453	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.900%
50) Toluene-d8	10.380	98	1051864	48.662	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.320%
62) 4-Bromofluorobenzene	12.668	95	328188	45.747	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082622\
Data File : VN074132.D
Acq On : 26 Aug 2022 13:12
Operator : JC\MD
Sample : VN0826WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0826WBL01

Quant Time: Aug 27 00:19:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 06:44:47 2022
Response via : Initial Calibration

