

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082622\
 Data File : VN074133.D
 Acq On : 26 Aug 2022 14:03
 Operator : JC\MD
 Sample : N4280-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE109D2-20220817

Quant Time: Aug 27 00:19:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 06:44:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	636068	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	954510	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	818125	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	354788	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	325607	52.557	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.120%
35) Dibromofluoromethane	7.951	113	286833	50.813	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.620%
50) Toluene-d8	10.380	98	1080010	49.401	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.800%
62) 4-Bromofluorobenzene	12.668	95	331870	45.740	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.480%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.145	101	2778	0.515	ug/l	88
27) cis-1,2-Dichloroethene	7.263	96	1864	0.273	ug/l	86
44) Trichloroethene	9.151	130	134433	18.809	ug/l	98
49) 1,4-Dioxane	9.516	88	1440	8.363	ug/l #	78
64) Tetrachloroethene	10.916	164	2760	0.427	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082622\
Data File : VN074133.D
Acq On : 26 Aug 2022 14:03
Operator : JC\MD
Sample : N4280-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE109D2-20220817

Quant Time: Aug 27 00:19:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 06:44:47 2022
Response via : Initial Calibration

