

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082622\
 Data File : VN074142.D
 Acq On : 26 Aug 2022 17:48
 Operator : JC\MD
 Sample : N4297-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW305I-20220820

Quant Time: Aug 27 00:21:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 06:44:47 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	387986	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	625901	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	548172	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	325190	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	219805	58.165	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.340%
35) Dibromofluoromethane	7.957	113	193586	52.299	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.600%
50) Toluene-d8	10.380	98	717827	50.073	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.140%
62) 4-Bromofluorobenzene	12.674	95	220174	46.277	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.560%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	4.122	96	1651	0.539	ug/l #	66
32) 1,1,1-Trichloroethane	7.939	97	2806	0.473	ug/l #	49
44) Trichloroethene	9.151	130	229508	48.970	ug/l	100
64) Tetrachloroethene	10.916	164	1839	0.425	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082622\
Data File : VN074142.D
Acq On : 26 Aug 2022 17:48
Operator : JC\MD
Sample : N4297-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW305I-20220820

Quant Time: Aug 27 00:21:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 06:44:47 2022
Response via : Initial Calibration

