

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082622\
 Data File : VN074143.D
 Acq On : 26 Aug 2022 18:13
 Operator : JC\MD
 Sample : N4297-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP03-20220820

Quant Time: Aug 27 00:22:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 06:44:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	575695	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	874860	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	741190	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	322413	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	305685	54.516	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.040%	
35) Dibromofluoromethane	7.957	113	265040	51.227	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.460%	
50) Toluene-d8	10.380	98	981315	48.973	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 97.940%	
62) 4-Bromofluorobenzene	12.669	95	304988	45.862	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.720%	
Target Compounds						
					Qvalue	
7) Trichlorofluoromethane	3.322	101	2802	0.333	ug/l	95
9) 1,1,2-Trichlorotrifluo...	4.134	101	1649	0.338	ug/l #	87
12) 1,1-Dichloroethene	4.122	96	2968	0.653	ug/l #	86
19) Methyl tert-butyl Ether	5.534	73	5856	0.358	ug/l	95
24) 1,1-Dichloroethane	6.316	63	3586	0.404	ug/l #	82
27) cis-1,2-Dichloroethene	7.245	96	3234	0.524	ug/l	67
44) Trichloroethene	9.151	130	2377400	362.910	ug/l	99
64) Tetrachloroethene	10.916	164	6402	1.094	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082622\
Data File : VN074143.D
Acq On : 26 Aug 2022 18:13
Operator : JC\MD
Sample : N4297-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP03-20220820

Quant Time: Aug 27 00:22:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 06:44:47 2022
Response via : Initial Calibration

