

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
 Data File : VN075553.D
 Acq On : 05 Dec 2022 12:46
 Operator : JC\MD
 Sample : VN1205MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1205MBL01

Quant Time: Dec 06 03:23:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	103267	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	170077	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	148521	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	60395	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	29131	52.552	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.100%
35) Dibromofluoromethane	8.171	113	35791	53.570	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.140%
50) Toluene-d8	10.571	98	68753	45.689	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.380%
62) 4-Bromofluorobenzene	12.853	95	48366	43.546	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
Data File : VN075553.D
Acq On : 05 Dec 2022 12:46
Operator : JC\MD
Sample : VN1205MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1205MBL01

Quant Time: Dec 06 03:23:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

