

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
 Data File : VN075554.D
 Acq On : 05 Dec 2022 13:10
 Operator : JC\MD
 Sample : VN1205WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1205WBL01

Quant Time: Dec 06 03:24:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	147034	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	258747	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	225308	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	92880	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	44986	56.998	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 114.000%	
35) Dibromofluoromethane	8.171	113	53921	53.049	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.100%	
50) Toluene-d8	10.570	98	108563	47.422	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 94.840%	
62) 4-Bromofluorobenzene	12.853	95	74930	44.344	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.680%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
Data File : VN075554.D
Acq On : 05 Dec 2022 13:10
Operator : JC\MD
Sample : VN1205WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBL01

Quant Time: Dec 06 03:24:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

