

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
 Data File : VN075557.D
 Acq On : 05 Dec 2022 14:31
 Operator : JC\MD
 Sample : N5818-04
 Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-113022

Quant Time: Dec 05 14:52:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	198236	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	319484	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	267303	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	114134	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	60171	56.546	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 113.100%	
35) Dibromofluoromethane	8.171	113	73615	58.656	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 117.320%	
50) Toluene-d8	10.565	98	157213	55.617	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.240%	
62) 4-Bromofluorobenzene	12.853	95	106446	51.019	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.040%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120522\
Data File : VN075557.D
Acq On : 05 Dec 2022 14:31
Operator : JC\MD
Sample : N5818-04
Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-04-113022

Quant Time: Dec 05 14:52:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

